



### ANN 03 Lab Equipment's

S.No	Department	Inst. ID	Machine Name	Make	Model no/Range.	Serial Number	Installation Date	Calibration			Done by
								frequency	Done	Due	
1	Biochemistry	BIO/001	Biochemistry Analyzer	Diasys	Respons 910	471135	3/9/2018	Yearly	4/25/2019	24.4.20	Vendor
2	Biochemistry	BIO/101	Immunocassay Analyzer	Roche	COBAS e 411	64D308	3/15/2018	NA	NA	NA	Vendor
3	Biochemistry	BIO/002	Electrolyte Analyzer	Transasia	Easylite plus	56330CNKC	3/6/2018	NA	NA	NA	Vendor
4	Hematology	HEM/101	Hematology Analyzer	Horiba	Pentra XL 80	601PXL7051	3/8/2018	Yearly	6/15/2019	14.6.20	Vendor
5	Hematology	HEM/102	Hematology Analyzer	Horiba	ES60	708ES0H1229	4/21/2018	Yearly	6/15/2019	14.6.20	Vendor
6	Hematology	HEM/001	Coagulation analyzer	stago	kc-1delta	17G63096	3/12/2018	Yearly	6/1/2019	31.5.20	Vendor
7	Hematology	HEM/003	Sample Roller	GEM	NA	A2025	3/19/2018	NA	NA	NA	Inhouse
8	Clinical Pathology	CLI/001	Urine Analyzer	Diasys	Ds 100	LD56471008	3/16/2018	Yearly	4/25/2019	24.4.20	Vendor
9	Clinical Pathology	CLI/002	Rotary Shaker	REMI	RS-12R	ZCHS-22995	3/19/2018	Half Yearly	6/1/2019	30.11.19	Inhouse
10	Accessioning	ACC/001	Centrifuge	REMI	R-8c plus	ZDHN-20665	3/19/2018	Half Yearly	6/1/2019	30.11.19	Inhouse
11	General	GEN/001	Centrifuge	REMI	R-8c plus	ZDHN-20669	3/19/2018	Half Yearly	6/1/2019	30.11.19	Inhouse
12	General	GEN/002	Microscope	Olympus	CH20i	17E0550	3/19/2018	NA	NA	NA	Inhouse
13	General	GEN/003	Incubator	International Scientific	NA	NA	3/19/2018	Half Yearly	6/1/2019	30.11.19	Inhouse
14	General	GEN/004	Refrigerator	Samsung	RT42M553ESL	0FST4DAJ801413D	3/19/2018	Half Yearly	6/1/2019	30.11.19	Inhouse
15	General	GEN/005	Refrigerator	Samsung	RT42M553ESL	0F5T4DAJ801443Y	3/19/2018	Half Yearly	6/1/2019	30.11.19	Inhouse
16	General	GEN/006	Refrigerator	Samsung	RT42M553ESL	0FST4DAJ900764F	3/19/2018	Half Yearly	6/1/2019	30.11.19	Inhouse