



ANN 04 Proficiency Testing

| Sl. no | Product/ Material | Details of Test(s)/ examination | Date of Testing/ examination | Organizing body | Performance in Standard deviation index | Corrective action taken (if required) |
|--------|--------------------|---------------------------------|------------------------------|-----------------|---|---|
| 1 | Lyophilized Sample | Albumin | 08.05.19 | RML - QAP | -3.0 | Equas sample rerun again. Result mentioned in CAPA logsheet |
| 2 | Lyophilized Sample | Alkaline phosphatase | 08.05.19 | RML - QAP | -1.7 | |
| 3 | Lyophilized sample | Bilirubin total | 08.05.19 | RML - QAP | 1.8 | |
| 4 | Lyophilized sample | Calcium total | 08.05.19 | RML - QAP | 1.2 | |
| 5 | Lyophilized sample | Cholesterol total | 08.05.19 | RML - QAP | -0.4 | |
| 6 | Lyophilized sample | Creatinine | 08.05.19 | RML - QAP | 4.3 | Kit changed |
| 8 | Lyophilized sample | Glucose | 08.05.19 | RML - QAP | -0.7 | |
| 9 | Lyophilized sample | HDL | 08.05.19 | RML - QAP | 0.6 | |
| 10 | Lyophilized sample | Potassium | 08.05.19 | RML - QAP | -1.5 | |
| 11 | Lyophilized sample | Total protein | 08.05.19 | RML - QAP | 0.3 | |
| 12 | Lyophilized sample | Sodium | 08.05.19 | RML - QAP | 0.1 | |
| 13 | Lyophilized sample | SGPT | 08.05.19 | RML - QAP | 0.6 | |
| 14 | Lyophilized sample | SGOT | 08.05.19 | RML - QAP | -0.8 | |

| | | | | | | |
|----|--------------------|--------------|----------|-----------|------|-------------|
| 15 | Lyophilized sample | Triglyceride | 08.05.19 | RML - QAP | 0.3 | |
| 17 | Lyophilized sample | Urea | 08.05.19 | RML - QAP | 3.2 | Kit changed |
| 18 | Lyophilized sample | Uric acid | 08.05.19 | RML - QAP | -0.8 | |

Participation in ISHTM - AIIMS, FEB 2019 , Hematology

| Sl. no | Product/ Material | Details of Test(s)/ examination | Date of Testing/ examination | Organizing body | Performance in Standard deviation index | Corrective action taken (if required) |
|--------|-------------------|---------------------------------|------------------------------|-----------------|---|---|
| 1 | | WBC | 08.05.19 | RML - QAP | 0.3 | |
| 2 | | RBC | 08.05.19 | RML - QAP | -0.05 | |
| 3 | | HB | 08.05.19 | RML - QAP | -0.39 | |
| 4 | | HCT | 08.05.19 | RML - QAP | -0.04 | |
| 5 | | MCV | 08.05.19 | RML - QAP | 0.03 | |
| 6 | | MCH | 08.05.19 | RML - QAP | -0.43 | |
| 8 | | MCHC | 08.05.19 | RML - QAP | -0.26 | |
| 9 | | Platelets | 08.05.19 | RML - QAP | 0.15 | |
| 10 | | Retic count | 08.05.19 | RML - QAP | 28.3 | Random error. Results will be monitored in subsequent EQAS cycles |