

CALIBRATION CERTIFICATE

CUSTOMER : BALAJI MEDICAL CENTER  
ADDRESS : NO 4/18,JAGADEESWARAN STREET,  
T.NAGAR,CHENNAI-600017

REPORT NO : 311

DATE OF CALIBRATION : 31-01-2020

MODEL : DIATRON

Name of equipment : ICOUNT5 5PART

CERTIFICATE NO : 311

NEXT CALIBRATION : 30-01-2021

SR-NO : 550039

CALIBRTATION DONE BY : MR.G.V.KARTHIK

PARAMETER	RANGE	ACCURACY
WBC.	0.1-99.9 X 10 <sup>3</sup> / $\mu$ l.	3%
RBC.	0.2-7.0 X 10 <sup>6</sup> / $\mu$ l.	2%
HGB.	0.1 - 25.0 g/dL.	2%
HCT.	$\leq$ 0.5 %.	3%
PLT	1-999 X 10 <sup>3</sup> / $\mu$ l	5%

We hereby the above equipment was calibrated and instrument was in good condition.

FOR JEEVAN HEALTHCARE SERVICES PVT LTD (OPC)

G.V.KARTHIK  
SERVICE MANA

