

Dr Dharap's Diagnostic Centre

MONTHLY HAEMATOLOGY

CYCLE 12 SAMPLE 12

Explanation of codes used in this report

R - Results removed due to reconstitution error
N - No result returned
C - Result corrected

Authorised by: Stephen Doherty, RIQAS Manager

Issue No: 1

Issue Date: 11/12/2019

Randox Laboratories Limited
55 Diamond Road
CRUMLIN BT29 4QY
Tel: +44 (0)28 9445 4399
Fax: +44 (0)28 9445 4398
Email: mail@riqas.com

Analyte	Mean for Comparison	Your Result	SDI	RMSDI	%DEV	RM%DEV	TS	RMTS	Performance
Haemoglobin	15.349	15.200	-0.39	-0.38	-1.0	-0.9	113	105	
Haematocrit (HCT)	43.085	43.300	0.13	0.58	0.5	2.1	120	96	
MCH	30.618	30.000	-0.65	-0.30	-2.0	-0.9	98	100	
MCHC	35.675	35.000	-0.51	-0.82	-1.9	-3.0	101	86	
MCV	86.020	85.000	-0.38	0.73	-1.2	2.3	120	93	
Mean Platelet Volume	9.326	9.400	0.12	0.02	0.8	0.1	120	113	
Platelets (Impedance Count)	271.899	276.000	0.20	0.16	1.5	1.2	120	110	
RBC (Impedance Count)	5.016	5.070	0.50	-0.12	1.1	-0.3	102	84	
Red Cell Dist. Width CV	13.661	14.300	0.82	0.49	4.7	2.8	80	73	
WBC (Impedance Count)	12.037	12.400	0.67	1.64	3.0	7.4	89	87	

ORMSDI 0.20

ORM%DEV 1.1

ORMTS 95

END OF REPORT