

Dr Dharap's Diagnostic Centre

MONTHLY CLINICAL CHEMISTRY

CYCLE 16 SAMPLE 12

Explanation of codes used in this report

- R - Results removed due to reconstitution error
- N - No result returned
- C - Result corrected

Authorised by: Stephen Doherty, RIQAS Manager

Issue No: 1

Issue Date: 02/01/2020

Randox Laboratories Limited
55 Diamond Road
CRUMLIN BT29 4QY
Tel: +44 (0)28 9445 4399
Fax: +44 (0)28 9445 4398
Email: mail@riqas.com

Analyte	Mean for Comparison	Your Result	SDI	RMSDI	%DEV	RM%DEV	TS	RMTS	Performance
Albumin	4.051	4.000	-0.22	-0.46	-1.3	-2.6	120	94	
Alkaline Phosphatase	200.036	220.000	0.60	0.99	10.0	12.4	73	61	
ALT (GPT)	39.467	40.500	0.30	0.89	2.6	7.9	120	82	
Amylase, Total	90.083	93.700	0.44	1.05	4.0	9.4	107	64	
AST (GOT)	34.847	32.900	-0.67	0.16	-5.6	1.3	89	103	
Bilirubin, Direct	0.848	0.740	-0.84	0.23	-12.7	3.5	79	87	
Bilirubin, Total	1.648	1.720	0.45	0.96	4.4	9.3	106	78	
Calcium	8.998	9.900	2.06	0.70	10.0	3.4	40	71	▲
Calcium, Ionised	0.959	1.110	1.11	Too Few	15.8	Too Few	65	Too Few	
Chloride	100.600	98.000	-0.92	-0.35	-2.6	-1.0	75	99	
Cholesterol	152.547	156.000	0.40	0.61	2.3	3.5	112	97	
CK, Total	196.882	221.100	1.67	1.14	12.3	9.0	42	74	
Creatinine	1.520	1.320	-1.75	-0.15	-13.2	-1.1	47	94	
GGT	45.866	No Result		Too Few		Too Few		Too Few	
Glucose	111.034	119.000	1.39	0.22	7.2	1.1	57	108	
HDL-Cholesterol	56.284	57.000	0.11	0.21	1.3	2.5	120	114	
LD (LDH)	410.236	399.900	-0.32	0.23	-2.5	1.8	120	87	
Lipase	36.585	No Result		Too Few		Too Few		Too Few	
Phosphate, Inorganic	4.487	4.220	-1.14	-0.83	-6.0	-4.3	66	52	
Potassium	3.831	3.800	-0.25	0.54	-0.8	1.7	120	95	
Protein, Total	5.820	5.270	-1.81	-0.94	-9.4	-4.9	46	78	
PSA, Total	16.258	17.680	0.58	-0.15	8.7	-2.3	91	85	
Sodium	138.300	139.000	0.24	0.08	0.5	0.2	120	91	
Total T3	1.447	1.380	-0.33	-0.32	-4.6	-4.7	119	106	
Total T4	7.395	7.470	0.08	0.11	1.0	1.5	120	109	
Trig Total	99.050	88.000	-1.38	-0.80	-11.2	-6.3	55	79	
TSH	1.054	0.881	-1.44	-0.11	-16.4	-1.3	56	110	
Urea	44.980	48.300	1.00	-0.72	7.4	-5.3	72	81	
Uric Acid (Urate)	6.012	5.400	-1.58	-0.24	-10.2	-1.5	52	81	

ORMSDI 0.12

ORM%DEV 1.3

ORMTS 88

END OF REPORT