

LAB MONTHLY SUMMARY

Lab Name **SRV PATH LABS PVT LTD** Lab No **11001**
 Month **December** Year **2019**
 Constituent Group **Chemistry I**

Date of Result Entered : 18/12/2019

Date of Report Published : 04/01/2020

Sl.No	Analyte	Method	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
1	GLUCOSE	GOD-POD	2411	423.19	9.13	38.65	383.3 mg/dl	-1.03	1.57
2	UREA	UREASE UV / GLDH	2441	137.28	10.79	14.81	156.1 mg/dl	1.27	0.60
3	CREATININE	JAFFES KINETIC	2392	7.98	11.12	0.89	10.8 mg/dl	3.18	0.04
4	T.BILIRUBIN	JENDRASSIK GROF	2368	2.42	16.89	0.41	2.4 mg/dl	-0.05	0.02
5	T-PROTEIN	BIURET	2783	5.16	8.91	0.46	5.4 g/dl	0.52	0.02
6	ALBUMIN	BCG	2742	3.15	8.54	0.27	3.4 g/dl	0.93	0.01
7	CALCIUM	ARSENazo III	1507	9.39	8.13	0.76	10.1 mg/dl	0.93	0.04
8	PHOSPHORUS	COLORIMETRIC WITHOUT PRECIPITATION	1321	3.80	12.69	0.48	4.1 mg/dl	0.62	0.03
9	URIC ACID	ENZYMATIC	2602	6.76	10.84	0.73	8.2 mg/dl	1.96	0.03
10	CHOLESTEROL	CHOD-PAP	2817	101.99	9.48	9.67	96 mg/dl	-0.62	0.36
11	TRIGLYCERIDE	ENZYMATIC	2618	83.87	18.58	15.58	94.3 mg/dl	0.67	0.61
12	AST	UV-Kinetic with PLP (P-5-P)	882	305.47	16.39	50.07	324.8 U/L	0.39	3.37
13	ALT	UV-Kinetic with PLP (P-5-P)	889	31.30	20.44	6.40	22.8 U/L	-1.33	0.43
14	ALP	PNP DEA KINETIC	258	306.54	20.70	63.45	409 U/L	1.61	7.90

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****