

## LAB MONTHLY SUMMARY

Lab Name DR.AMRISH MEHTA DIAGNOSTIC CENTRE Lab No 0281  
 Month October Year 2019  
 Constituent HbA1c  
 Group

Date of Result Entered : 19/10/2019

Date of Report Published : 01/11/2019

SI.No	Analyte	Method	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
1	HbA1c	HPLC-ION EXCHANGE	543	7.27	5.87	0.43	7 %	-0.63	0.04

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*



1/11/19