

DR. DASGUPTA'S PATHOLOGICAL LABORATORY

F-26/10, SECTOR-7, ROHINI, DELHI - 110085
PHONE : 47090336

VIEW LAB MONTHLY SUMMARY

Lab Name DR DASGUPTA S PATHOLOGICAL LABORATORY
Lab No 11322
Month January
Year 2020
Constituent Group Chemistry I

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Method Result

Print

Date of Result Entered : 20/01/2020

Date of Report Published : 11/02/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE	GOD-POD	Transasia / Erba	760	128.80	7.39	9.52	152 mg/dl	2.44	0.69
2	UREA	UREASE UV / GLDH	Transasia / Erba	677	41.02	8.75	3.59	47 mg/dl	1.67	0.28
3	CREATININE	JAFFE-END POINT	Transasia / Erba	207	1.80	9.21	0.17	1.7 mg/dl	-0.60	0.02
4	T.BILIRUBIN	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Transasia / Erba	676	1.66	17.84	0.30	1.9 mg/dl	0.81	0.02
5	T-PROTEIN	BIURET - colorimetric	Transasia / Erba	726	5.48	7.17	0.39	6.2 g/dl	1.83	0.03
6	ALBUMIN	BCG - colorimetric	Transasia / Erba	698	3.28	7.88	0.26	3.3 g/dl	0.08	0.02
7	CALCIUM	ARSENazo III	Transasia / Erba	457	8.79	8.87	0.78	10 mg/dl	1.55	0.07
8	PHOSPHORUS	Molybdate UV/ Phosphomolybdate complex	Transasia / Erba	299	5.02	11.77	0.59	4.8 mg/dl	-0.37	0.07
9	URIC ACID	ENZYMATIC / URICASE Colorimetric	Transasia / Erba	664	5.63	10.34	0.58	6.6 mg/dl	1.67	0.05
10	CHOLESTEROL	CHOD-PAP	Transasia / Erba	734	109.32	9.79	10.71	164 mg/dl	5.11	0.79
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	639	75.70	19.61	14.85	95 mg/dl	1.30	1.17
12	AST	UV-Kinetic with PLP (P-5-P)	Transasia / Erba	235	73.07	14.01	10.24	78 U/L	0.48	1.34
13	ALT	UV-Kinetic with PLP (P-5-P)	Transasia / Erba	243	30.86	20.76	6.40	35 U/L	0.65	0.82
14	ALP	PNP AMP KINETIC	Transasia / Erba	583	82.85	14.43	11.96	85 U/L	0.18	0.99
15	AMYLASE	CNPG3	Biosystems	197	45.47	9.89	4.50	35.3 U/L	-2.26	0.64

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

This is a professional opinion and not the final diagnosis. It should be clinically correlated.
home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=225&noSamples=1&month=1&year=2020&groupId=1&yea... 1/2