



Certificate No. CC-2267

# FINIKS ENGINEERING COMPANY

Plot No. 21, Anand Nagar, Near Wasan Automobiles, Sahyadri Chowk, M.I.D.C.,  
A'nagar - 414111. [M.S.] Tel. [0241] 2778918, E-mail : finiks97@rediffmail.com

## Calibration Certificate

ULR : CC22672010000499F

CALIBRATED ON	CERTIFICATE NO	NEXT CALIBRATION DUE ON	ISSUE DATE	PAGE
21/01/2020	20-SMAC21/01/001	21/01/2021	29/01/2020	1/1

**CALIBRATED FOR** : M/s. Dr. Amrish Mehta's Apple Diagnostic  
2nd Floor Prabhu Krupa Bldg , Tilak Road  
Opp Silver Solitaire Bldg , Next To Bank of Baroda ,  
Ghatkopar ( East ) , Mumbai 400077

**CUSTOMER'S CHALLAN NO.** : On Site Work order No.: 20-SMA21/01

**TYPE OF INSTRUMENT** : CENTRIFUGE ( RPM )

**IDENTIFICATION**

Make	Remi / R-8C Plus	Sr No	ZEAN-03612
Range	0 - 6000 RPM	Id. No.	--
L.C.	1 RPM	Dept	Bio- Chemistry / Immunology
Location	Lab		

**LAB ID NO.** : 20-SMA21/01/001 **Condition at Receipt** : OK

**ENVIRONMENTAL CONDITION** : Temperature : 24 °C Humidity : 54 %

**REFERENCE DOCUMENT USED** : IS 12508 - 1988 & Sanas TR 45-01 (2008-2-15)

**CALIBRATION PROCEDURE** : CP/81/00 **Calibration Location** : Site

**Equipment / Reference Standards used for Calibration** : Traceable to National/International Standards

Equipment / Reference Standard	Identification No.	Certificate No.	Calib. Date	Valid up to
Tachometer	Finiks/TM-01	40012543	25/03/2019	24/03/2020


### CALIBRATION RESULTS

Parameter : RPM

Calibration Point	STD Reading	UUC Reading	Error
RPM	RPM	RPM	RPM
1500	1504	1500	-4.0
2000	2003	2000	-3.0
2500	2504	2500	-4.0
3000	3007	3000	-7.0
3500	3506	3500	-6.0

Uncertainty of Measurement : ± 4.76 RPM

The reported uncertainty is at coverage factor K=2 which corresponds to a coverage probability of approximately 95% for normal distribution.



**Calibrated by** *[Signature]* **Authorized Signatory** *[Signature]* **Warpe B. S.**  
**T. M.**

This Certificate refers only to the item (s) submitted for calibration  
This calibration results in this certificate are valid at the time of and under the stated conditions of measurements.  
This certificate shall not be reproduced, except in full without our prior permission in writing

FM/17/00