



Certificate No. CC-2267

# FINIKS ENGINEERING COMPANY

Plot No. 21, Anand Nagar, Near Wasan Automobiles, Sahyadri Chowk, M.I.D.C.,  
A'nagar - 414111. [M.S.] Tel. [0241] 2778918, E-mail : finiks97@rediffmail.com

## Calibration Certificate

ULR:CC22672000002434F

CALIBRATED ON	ISSUED ON	CERTIFICATE NO	NEXT CALIBRATION DUE ON	PAGE	NO. OF PAGES
20/01/2020	29/01/2020	20-MAC-20-01-001	20/01/2021	1	1

**CALIBRATED FOR** : **M/s Dr.Amrish Mehta's Apple Diagnostic.**  
2nd Floor Prabhu Krupa bldg, Tilak Road ,  
Opp Silver Solitaire bldg, Next to Bank of Baroda  
Ghatkopar (East), Mumbai 400077.

Date of Receipt: 20/01/2020

**CUSTOMER'S CHALLAN NO.** :--

Work Order No: 20-MA-20-01

**TYPE OF INSTRUMENT:** MICROPIPETTE  
**IDENTIFICATION**  
Range : 100 to 1000µl      **Make** : Thermo Scientific  
L.C : 1µl      **Sr.No.** :--  
Model : Finn pipette F3      **Id.No.** : HW04164  
Location : Biochem, Hematology, Serology, Immunology

**LAB ID NO.** : 20-MA-20-01-001      **Condition at time of calibration** : OK

**ENVIRONMENTAL CONDITION :**

Temp.	26 °C	Humidity	54 %Rh	Barometric Pressure	935.4mbar
-------	-------	----------	--------	---------------------	-----------

**REFERENCE DOCUMENT USED:** ISO 8655

**CALIBRATION PROCEDURE** : CP/101/00      Calibration Location: At Lab

**Equipment / Reference Standards used for Calibration** : Traceable to National/International Standards

Equipment / Reference Standard	Identification No.	Certificate No.	Calib. Date	Valid up to
Digital Weighing Balance	Finiks/DWB-01	NCQC-MO/290319/01	29/03/2019	28/03/2021

### CALIBRATION RESULTS

Sr.No	UUC Reading in (µl)	Observed Reading in water (µl)	Exp.Uncertainty (±)K=2
1	100.0	99.67	1.39µl
2	500.0	482.76	1.39µl
3	1000.0	996.87	1.39µl

The reported uncertainty is at coverage factor K=2 which corresponds to a coverage probability of approximately 95% for normal distribution

**Calibrated by** *Mokal*      **Authorized Signatory** *Warda D.S.*

This Certificate refers only to the item (s) submitted for calibration.  
This calibration results in this certificate are valid at the time of and under the stated conditions of measurements.  
This certificate shall not be reproduced, except in full without our prior permission in writing

T.M.

FM/17/00

