



The Ultimate Calibration Services

The Quality Solutions...



Certificate No : CC-3020

Gala No. 1&2, Kamalkunj CHS, Opp. to Corporation Bank Lane, Near Kaka Gole Foundation, Kulgaon, Badlapur (E) Dist. Thane 421503
Phone No : 9890715627 | 9665003365 • Mail: theucs20@gmail.com • GST No. : 27AANFT7678Q1ZT

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/20010035

Date of Issue: 25-01-2020

Date of calibration: 19-01-2020

Suggested Due Date: 19-01-2021

ULR No:CC302020000000035F

Name and Address of Customer:
Dr. Amrish Mehta's Apple Diagnostics,
2 nd Floor, Prabhu Nagar Building, Tilak Road, Opp. Silver Solitaire Bldg.
Next to Bank of Baroda, Ghatkopar (East), Mumbai.

Ref. No: Letter
Date: 18-01-2020
Date of Receipt: 18-01-2020
Condition on Receipt: Good

Details of Equipment :

Name : Digital Thermohygrometer

Make : Mextech

Model: TM-01

Sr.No. : --

Id.No: ---

Department : Pathology

Location : ---

Least Count: 0.1 °C / 1 % RH

Range : 0 to 50 °C / 0 to 100 %RH

Accuracy : Not Specified

Calibration Points: 10, 25 & 35 °C ; 40, 50 & 60 %RH

Calibration Procedure Used: UCS/CAL/306

Environmental Conditions : 25 ± 3 °C & 50 ± 15 % RH

Details of STD equipment used : UCS/CAL/004

Name	Id No	Parameter/Uncertainty	Certificate No	Valid up to
Relative Humidity & Temperature Meter with Sensor	UCS/CAL/004	Temperature : ± 0.46°C Relative Humidity: ± 1.67 % RH	PICAL/1119/T/052 PI Cal Lab	13-11-2020

OBSERVATIONS :

Discipline : Thermal

Parameter : Temperature

Sub Group : Temperature

Range	Reading On UUC	Reading on Standard	Error	Expanded Uncertainty
°C	°C	°C	°C	°C
0 to 50	10.2	9.93	0.27	± 0.77
	25.2	25.26	-0.06	± 0.77
	35.9	35.63	0.27	± 0.77

Discipline : Thermal

Parameter : Humidity

Sub Group : Specific Heat & Humidity

Range	Reading On UUC	Reading on Standard	Error	Expanded Uncertainty
%RH	%RH	%RH	%RH	%RH
0 to 100	38	38.71	-0.71	± 2.7
	48	48.66	-0.66	± 2.7
	58	58.73	-0.73	± 2.7

Calibrated By : *S.Sawant*
Name & Sign: Sarvesh Sawant



Authorised Signatory : *V.W.Patwardhan*
Name & Designation: V.W.Patwardhan
Technical Manager

Format No: UCS/QF/7.8/01

Page no 1 of 2 P.T.O.



The Ultimate Calibration Services

The Quality Solutions...



Certificate No : CC-3020

Gala No. 1&2, Kamalkunj CHS, Opp. to Corporation Bank Lane, Near Kaka Gole Foundation, Kulgaon, Badlapur (E) Dist. Thane 421503
Phone No : 9890715627 | 9665003365 • Mail: theucs20@gmail.com • GST No. : 27AANFT7678Q1ZT

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/20010035

Date of Issue: 25-01-2020

Date of calibration: 19-01-2020

Suggested Due Date: 19-01-2021

ULR No:CC302020000000035F

Remarks:

1. The readings noted above are average of five readings.
2. The uncertainty expressed as Expanded uncertainty is calculated at the coverage factor $k=2$ at a confidence level of 95%.

The readings noted are particularly for the equipment submitted for calibration.

The reported values are valid at the time of calibration and under the controlled conditions.

Certificate should not be reproduced except in full with the prior approval of Laboratory.

Any handmade corrections made on the certificate, invalidates the Calibration Certificate.


The Ultimate Calibration Services will not be responsible for any mishandling of equipment after issueing the calibration certificate.

All precautions are being taken while handling the equipment at the time of calibration but lab is not liable for any misbehave of the equipment during calibration or any loss in the process of calibration.

The measurements are traceable to National/International standard.

Calibrated By :

Name & Sign : Sarvesh Sawant



Authorised Signatory :

Name & Designation: V.W.Patwardhan

Technical Manager

Format No: UCS/QF/7.8/01

-----End of Certificate-----

Page no

2 of 2