



The Ultimate Calibration Services

The Quality Solutions...



Certificate No : CC-3020

Gala No. 1&2, Kamalkunj CHS, Opp. to Corporation Bank Lane, Near Kaka Gole Foundation, Kulgaon, Badlapur (E) Dist. Thane 421503
Phone No : 9890715627 | 9665003365 • Mail: theucs20@gmail.com • GST No. : 27AANFT7678Q1ZT

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/20010033

Date of Issue: 25-01-2020

Date of calibration: 19-01-2020

Suggested Due Date: 19-01-2021

ULR No:CC30202000000033F

Name and Address of Customer:

Dr. Amrish Mehta's Apple Diagnostics,
2 nd Floor, Prabhu Nagar Building, Tilak Road, Opp. Silver Solitaire Bldg.
Next to Bank of Baroda, Ghatkopar (East), Mumbai.

Ref. No: Letter

Date: 18-01-2020

Date of Receipt: 18-01-2020

Condition on Receipt: Good

Details of Equipment :

Name : Digital Thermometer

Make : Multi

Model: --

Sr.No. : ---

Id.No: BR/DT/01

Department : Pathology

Location : Biochem, Hematology

Least Count: 0.1 °C

Range : -50 to 300 °C

Accuracy : Not Specified

Calibration Points: -20, 0,50,100 & 250 °C

Calibration Procedure Used: UCS/CAL/301

Environmental Conditions : 25 ± 3 °C & 50 ± 15 % RH

Details of STD equipment used : UCS/CAL/003, UCS/CAL/001

Name	Id No	Parameter/Uncertainty	Certificate No	Valid up to
Digital Multimeter	UCS/CAL/003	Resistance: ± 0.0012 %	M-191130-13-1 Godrej & Boyce	02-12-2020
RTD Sensor	UCS/CAL/001	Temperature : ± 0.35 °C	M-191130-13-2 Godrej & Boyce	30-11-2020


OBSERVATIONS :

Discipline : Thermal


Parameter : Temperature

Sub Group : Temperature

Range	Reading On UUC	Reading on Standard	Error	Expanded Uncertainty
°C	°C	°C		°C
-50 to 300	-20.3	-20.12	-0.18	± 0.41
	-1.1	-0.48	-0.62	± 0.41
	49.0	50.15	-1.15	± 0.70
	95.3	100.24	-4.94	± 0.70
	255.8	251.31	4.49	± 0.78

Calibrated By : 
Name & Sign: Sarvesh Sawant



Authorised Signatory : 
Name & Designation: V.W.Patwardhan
Technical Manager

Format No: UCS/QF/7.8/01

Page no 1 of 2



The Ultimate Calibration Services

The Quality Solutions...



Certificate No : CC-3020

Gala No. 1&2, Kamalkunj CHS, Opp. to Corporation Bank Lane, Near Kaka Gole Foundation, Kulgaon, Badlapur (E) Dist. Thane 421503
Phone No : 9890715627 | 9665003365 • Mail: theucs20@gmail.com • GST No. : 27AANFT7678Q1ZT

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/20010033

Date of Issue: 25-01-2020

Date of calibration: 19-01-2020

Suggested Due Date: 19-01-2021

ULR No:CC302020000000033F

Remarks:

1. The readings noted above are average of five readings.
2. The uncertainty expressed as Expanded uncertainty is calculated at the coverage factor $k=2$ at a confidence level of 95%.
3. Diameter of Probe = 3.78 mm & Total Length of Probe = 87.3 mm

The readings noted are particularly for the equipment submitted for calibration.

The reported values are valid at the time of calibration and under the controlled conditions.

Certificate should not be reproduced except in full with the prior approval of Laboratory.

Any handmade corrections made on the certificate, invalidates the Calibration Certificate.


The Ultimate Calibration Services will not be responsible for any mishandling of equipment after issuing the calibration certificate.

All precautions are being taken while handling the equipment at the time of calibration but lab is not liable for any misbehave of the equipment during calibration or any loss in the process of calibration.

The measurements are traceable to National/International standard.

Calibrated By : 
Name & Sign : Sarvesh Sawant



Authorised Signatory : 
Name & Designation: V.W. Patwardhan
Technical Manager

Format No: UCS/QF/7.8/01

-----End of Certificate-----

Page no

2 of 2