



Certificate No. CC-2267

FINIKS ENGINEERING COMPANY

Plot No. 21, Anand Nagar, Near Wasan Automobiles, Sahyadri Chowk, M.I.D.C.,
A'nagar - 414111. [M.S.] Tel. [0241] 2778918. E-mail : finiks97@rediffmail.com

Calibration Certificate

ULR : CC226720100000500F

CALIBRATED ON	CERTIFICATE NO	NEXT CALIBRATION DUE ON	ISSUE DATE	PAGE
21/01/2020	20-SMAC21/01/002	21/01/2021	29/01/2020	1/1

CALIBRATED FOR

: M/s. Dr. Amrish Mehta's Apple Diagnostic
2nd Floor Prabhu Krupa Bldg , Tilak Road
Opp Silver Solitaire Bldg , Next To Bank of Baroda ,
Ghatkopar (East) , Mumbai 400077

CUSTOMER'S CHALLAN NO.

: On Site Work order No.: 20-SMA21/01

TYPE OF INSTRUMENT

: CENTRIFUGE (RPM)

IDENTIFICATION

Make	Remi / R-8C	Sr No	ZKN-32414
Range	0 - 6000 RPM	Id. No.	--
L.C.	1 RPM	Dept	Clinical Pathology
Location	Lab		

LAB ID NO.

: 20-SMA21/01/002 Condition at Receipt : OK

ENVIRONMENTAL CONDITION

: Temperature : 24 °C Humidity : 54 %

REFERENCE DOCUMENT USED

: IS 12508 - 1988 & Sanas TR 45-01 (2008-2-15)

CALIBRATION PROCEDURE

: CP/81/00 Calibration Location : Site

Equipment / Reference Standards used for Calibration : Traceable to National/International Standards

Equipment / Reference Standard	Identification No.	Certificate No.	Calib. Date	Valid up to
Tachometer	Finiks/TM-01	40012543	25/03/2019	24/03/2020

CALIBRATION RESULTS

Parameter : RPM

Calibration Point	STD Reading	UUC Reading	Error
RPM	RPM	RPM	RPM
1500	1526	1524	-2.0
2000	2014	2016	2.0
2500	2502	2502	0.0
3000	2995	2995	0.0
3300	3288	3296	8.0

Uncertainty of Measurement : ± 4.76 RPM

The reported uncertainty is at coverage factor K=2 which corresponds to a coverage probability of approximately 95% for normal distribution.

Calibrated by <i>[Signature]</i>	Authorized Signatory <i>[Signature]</i> Warpe B. S. T. M.
----------------------------------	--

This Certificate refers only to the item (s) submitted for calibration
This calibration results in this certificate are valid at the time of and under the stated conditions of measurements.
This certificate shall not be reproduced, except in full without our prior permission in writing

FM/17/00