

MANGALORE EQAP REPORT
NOVEMBER- 2019

<u>SL. NO.</u>	<u>HEAMATOLOGY</u>	<u>VALUE-1</u>	<u>HEAMATOLOY</u>	<u>VALUE-2</u>
1	WBC	5.3	WBC	5.5
2	RBC	4.32	RBC	4.38
3	HGB	13.5	HGB	13.6
4	HCT	44.1	HCT	44.5
5	MCV	102.1	MCV	101.8
6	MCH	31.2	MCH	31.1
7	MCHC	30.5	MCHC	30.6
8	PLT	1.95	PLT	1.97

DATE : 05/11/2019 TIME : 2.55PM

PREPARED BY: MS. SOWMYA/DEEPIKA .J





PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 472

Distribution No.: 149-B

Month/Year: November/2019

Instrument ID: I COUNT 5

Name & Contact No. of PT Co-ordinator: Dr. Renu Saxena (Prof & Head), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 12-12-2019[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	5.5	5.3	10.8	8.9	0.0230	2.77	0.2	0.1	0.0050	1.25
RBC x10 ⁶ /µl	1	4.4	4.3	8.7	8.26	0.0070	2.00	0.1	0.04	0.0010	2.19
Hb g/dl	1	13.6	13.5	27.1	27.4	0.0240	-0.45	0.1	0.1	0.0060	0.00
HCT%	1	44.5	44.1	88.6	83	0.1870	0.79	0.4	0.4	0.0200	0.00
MCV-fl	1	102.1	101.8	203.9	199.8	0.4020	0.30	0.3	0.2	0.0160	0.34
MCH-Pg	1	31.2	31.1	62.3	66.6	0.0540	-2.76	0.1	0.2	0.0160	-0.45
MCHC-g/dl	1	30.6	30.5	61.1	66.2	0.1580	-0.96	0.1	0.2	0.0100	-0.34
Plt. x10 ³ /µl	1	197	195	392	387	2.02	0.16	2	5	0.27	-0.58
Retic %	2	5.9	5.3	11.2	9	0.21	0.33	0.6	0.25	0.02	1.18

P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=02/100WBC , Poly=--- L=06%, E=---, Mono/Promono=---, B1=90% P.M.=---, Mye=---, Meta=---, Other=---
RBC Morphology	3	NORMOCYTIC NORMOCHROMIC
Diagnosis	3	ACUTE LEUKEMIA(MORPHOLOGICALLY ACUTE MYELOID LEUKEMIA)

Blast: 80-90, Poly: 5-7, Lymph: 2-7, Eo/Mono/Pro/My/Meta: 0-4

Predominantly: Normocytic Normochromic. Moderate: Anisocytosis. Mild: Microcytic

Acute Leukemia (myeloid lineage)

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 ³ /µl	1	450	388	88.92	90.72	6.96	1.29	3.61	7.47
RBC x10 ⁶ /µl	1	450	388	91.49	88.4	4.38	4.64	3.61	5.93
Hb g/dl	1	450	388	90.21	89.18	4.64	4.9	4.64	5.41
HCT%	1	450	388	97.68	92.01	1.8	3.09	0	4.12
MCV-fl	1	450	388	95.88	84.54	2.58	10.31	0.77	4.38
MCH-Pg	1	450	388	87.89	90.98	5.67	3.61	5.41	4.38
MCHC-g/dl	1	450	388	95.88	89.95	2.32	4.38	1.03	4.64
Plt. x10 ³ /µl	1	450	388	92.53	86.86	5.15	6.19	1.8	6.44
ReticCount%	2	450	305	92.46	83.28	3.61	2.62	3.28	15.74
PS Assessment	3	450	363	Acceptable:94.9%,Warning Signal:2.7%,Unacceptable :2.4%					

Comments:

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Report authorized by,



Dr. R. Saxena

Prof & Head, Hematology, AIIMS, Delhi.

PT Co-ordinator: ISHTM-AIIMS-EQAP

-----End Of Report-----