

**CALIBRATION CERTIFICATE**

CUSTOMER : BALAJI MEDICAL CENTER  
ADDRESS : SHOP NO 5, DOOR#1-65/31,  
KULLOOR-KAVOOR ROAD, PADUKODI, PANJIMOGARU, MANGALORE-5750132  
REPORT NO : 911 CERTIFICATE NO : 911  
DATE OF CALIBRATION : 05-03-2019 NEXT CALIBRATION : 04-03-2020  
MODEL : DIATRON  
Name of equipment : ICOUNT5 SPART SR-NO : 550033  
CALIBRATION DONE BY : MR.G.V.KARTHIK

PARAMETER	RANGE	ACCURACY
WBC.	0.1-99.9 X 10 <sup>3</sup> / $\mu$ l.	3%
RBC.	0.2-7.0 X 10 <sup>6</sup> / $\mu$ l.	2%
HGB.	0.1 - 25.0 g/dL.	2%
HCT.	$\leq$ 0.5 %.	3%
PLT	1-999 X 10 <sup>3</sup> / $\mu$ l	5%

We hereby the above equipment was calibrated and instrument was in good condition.

FOR JEEVAN HEALTHCARE SERVICES PVT LTD (OPC)

  
G.V.KARTHIK  
SERVICE MANAGER