



**Jeevan Healthcare**  
Services Private Limited (OPC)

## CALIBRATION CERTIFICATE ✓

CUSTOMER : BALAJI MEDICAL CENTER  
ADDRESS : 38, Manikodi Srinivasan nagar Main Road Rajiv Gandhi salai,  
perungudi Chennai -600096

REPORT NO : 1921  
DATE OF CALIBRATION : 20-11-2019  
MODEL : DIATRON  
Name of equipment : ICOUNT5 5PART

CERTIFICATE NO : 1921  
NEXT CALIBRATION : 19-11-2020  
SR-NO : 550025

CALIBRATION DONE BY : MR.G.V.KARTHIK

PARAMETER	RANGE	ACCURACY
WBC.	0.1-99.9 X 10 <sup>3</sup> / $\mu$ l.	3%
RBC.	0.2-7.0 X 10 <sup>6</sup> / $\mu$ l.	2%
HGB.	0.1 - 25.0 g/dL.	2%
HCT.	≤ 0.5 %.	3%
PLT	1-999 X 10 <sup>3</sup> / $\mu$ l	5%

We hereby the above equipment was calibrated and instrument was in good condition.



G.V. KARTHIK  
FOR JEEVAN HEALTHCARE SERVICES PVT LTD (OPC)

G.V. KARTHIK  
SERVICE MANAGER

Jeevan Healthcare Services Private Limited (OPC)

📍 No. 12, 72<sup>nd</sup> Street, R V Nagar,  
Ashok Nagar, Chennai - 600 083.

☎ 044 - 45546464 / +91 - 9940189123

✉ info@jeevanhealthcare.in

🌐 www.jeevanhealthcare.in