

2/13/2020

LAB MONTHLY SUMMARY

Lab Name **BALAJI MEDICAL CENTRE** Lab No **5817**
 Month **January** Year **2020**
 Constituent Group **Chemistry II**

Date of Result Entered : 13/01/2020

Date of Report Published : 11/02/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	OTHERS (Any other Analyzer)	850	134.43	9.69	13.03	142 mg/dl	0.58	0.89
2	UREA I	UREASE BERTHELOT	OTHERS (Any other Analyzer)	281	39.56	14.36	5.68	36.8 mg/dl	-0.49	0.68
3	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) / JENDRASSIK	OTHERS (Any other Analyzer)	597	2.04	21.38	0.44	2 mg/dl	-0.09	0.04
4	T-PROTEIN I	BIURET - colorimetric	OTHERS (Any other Analyzer)	690	5.22	11.48	0.60	5.3 g/dl	0.13	0.05
5	ALBUMIN I	BCG - colorimetric	OTHERS (Any other Analyzer)	674	3.19	10.36	0.33	3.01 g/dl	-0.54	0.03
6	URIC ACID I	ENZYMATIC / URICASE Colorimetric	OTHERS (Any other Analyzer)	631	5.96	15.93	0.95	5.9 mg/dl	-0.06	0.08
7	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	OTHERS (Any other Analyzer)	766	115.15	15.91	18.32	108.9 mg/dl	-0.34	1.32

Perfect

All are within -1.0 to 1.0 (Excellent)

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

2/13/2020

LAB MONTHLY SUMMARY

Lab Name BALAJI MEDICAL CENTRE Lab No 5817
 Month January Year 2020
 Constituent Group HbA1c

Date of Result Entered : 13/01/2020

Date of Report Published : 11/02/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	HbA1c	TURBIDIMETRIC INHIBITION IMMUNO ASSAY	OTHERS (Any other Analyzer)	171	5.33	11.42	0.61	5.3 %	-0.05	0.09

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****