



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 1597

Distribution No.: 149-C

Month/Year: November/2019

Instrument ID: I COUNT 5

Name & Contact No. of PT Co-ordinator: Dr. Renu Saxena (Prof & Head), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 24-12-2019[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	17.4	17.2	34.6	30.11	0.0150	3.19	0.2	0.2	0.0050	0.00
RBC x10 <sup>6</sup> /µl	1	3.2	3.2	6.4	6.74	0.0070	-2.08	0	0.03	0.0010	-0.81
Hb g/dl	1	11.7	11.5	23.2	23.3	0.0180	-0.17	0.2	0.1	0.0060	0.67
HCT%	1	37.1	36.4	73.5	74.05	0.1400	-0.13	0.7	0.3	0.0170	1.08
MCV-fl	1	113.8	113.2	227	219	0.3040	0.77	0.6	0.3	0.0180	0.81
MCH-Pg	1	35.8	35.7	71.5	69.2	0.0530	1.24	0.1	0.3	0.0110	-0.67
MCHC-g/dl	1	31.6	31.5	63.1	63.1	0.1270	0.00	0.1	0.3	0.0110	-0.67
Plt. x10 <sup>3</sup> /µl	1	342	339	681	823	1.05	-2.20	3	8	0.30	-0.66
Retic %	2	5.4	5.1	10.5	7.2	0.16	0.65	0.3	0.4	0.02	-0.34

*All are perfect*

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT
DLC%	3	Nrbc=--- , Poly=stab-10% L=01%, E=01%, Mono/Promono=01% , B1=02% P.M.=04%, Mye=09%, Meta=07%, Other=---	Poly: 65-75, nRBC/Lymph/Eo/Mono/blast/pro: 0-5, Myelo: 10-15, Meta: 5-12, Baso: 0-2
RBC Morphology	3	NO SIGNIFICANT POLYCHROMASIA	Predominantly: Normocytic Normochromic, Moderate: Anisocytosis, Mild: Microcytic
Diagnosis	3	chronic myeloid leukemia-chronic phase	Chronic Myeloid Leukemia (Chronic Phase) [ CML-CP ]

*Rps*

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC $\times 10^3/\mu\text{l}$	1	450	373	89.81	93.03	4.29	3.22	5.63	3.49
RBC $\times 10^6/\mu\text{l}$	1	450	373	87.67	89.01	6.7	3.75	5.36	6.7
Hb g/dl	1	450	373	86.33	88.74	6.97	3.22	6.43	7.77
HCT%	1	450	373	96.51	89.54	1.34	4.29	1.88	5.09
MCV-fl	1	450	373	94.1	94.37	3.49	1.07	2.14	4.29
MCH-Pg	1	450	373	88.74	91.96	4.02	3.75	6.97	4.02
MCHC-g/dl	1	450	373	92.23	87.4	5.09	5.63	2.41	6.17
Plt. $\times 10^3/\mu\text{l}$	1	450	373	86.6	90.35	6.7	3.49	6.43	5.9
ReticCount%	2	450	266	90.23	95.49	4.89	0.38	3.01	6.77
PS Assessment	3	450	354	Acceptable:97%,Warning Signal:2.2%,Unacceptable :0.8%					

**Comments:**

- 1). Among Lab (EQA) : CBC result for WBC unacceptable, may be due to random/human error
- 2). Within Lab (IQA) : Precision acceptable.

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).


**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

Report authorized by,



Dr. R. Saxena

Prof & Head, Hematology, AIIMS, Delhi.

PT Co-ordinator: ISHTM-AIIMS-EQAP

-----End Of Report-----