

INSTALLATION REPORT

HIN-101- 5416

Customer Name : SHRADDDHADEEP GREEN CROSS PATHO. LABOR.
(Entity Name)
Address: PLOT NO. 910/1,
SECTOR 7-C, Pin Code: 382007 City: GANDHII-
MAGAR State: GUJARAT
OPP. CIVIL HOSPITAL, NS. GH 3 CIRCLE,
Contact Person: DR. DINESH RATHOD
(Dr./Lab Incharge)
Phone: 079 - 23222300 Mobile: E-mail:

INSTRUMENT & INVOICE:

Instrument Type: HAEMATOLOGY ANALYZER Date of Installation: 17/1/2020
Model: YH550 Date of Invoice:
Instrument Serial No.: 909YAXH02687 Invoice No :
Version: 1.0.4.2 Monitor S/N :
Printers S/N: CM34NSGIKS - HP INK TANK 419 Others :
CPU S/N :

PHYSICAL INSPECTION:

1. Was there any external Damage to the Instrument or Accessories ? Yes / No ✓
2. Were any of the accessories missing ? Yes / No ✓
3. Were any parts loose ? Any unsecured hardware ? Yes / No ✓
If any of the above were answered YES, please elaborate.
.....
.....

PRE-INSTALLATION CHECKS:

Is Room Air-Conditioned ? ✓ Yes / No
Vibration Free Platform ? ✓ Yes / No
Mains A.C. Supply :
Specified Capacity & Make : C.V.T. U.P.S. Stabilizer
Specified Capacity & Make : 2 KV
Output Voltage : Line & Neutral 220 Neutral & Earth 1V Line & Earth 230

TO BE COMPLETED BY THE CUSTOMER:

The above mentioned instrument has been satisfactorily installed by the Engineer / Product Specialist of Horiba India Pvt. Ltd
Received all standard accessories in good condition.
Customer's Remarks
.....
Signature of Customer : Name & Signature of Engineer :
Seal & Date : Date & Place : 17/1/20, GANDHINAGAR