



Dr. Amrish Mehta's, Apple Diagnostic, Prabh
Krupa, 2nd floor, Tilak road, Near Universal school,
Ghatkopar (East), Mumbai-400077

CBC+ESR ILC

Patient name: Ritu Sajnani

| Parameters | Apple lab Results | Aspira Lab Results | Nucleus Lab Results | CV% |
|------------|-------------------|--------------------|---------------------|------|
| Hb | 12.6 | 12.6 | 12.5 | 0.46 |
| PCV | 39.6 | 41 | 40 | 1.79 |
| RBC | 4.5 | 4.5 | 4.54 | 0.51 |
| ESR | 41 | 44 | 43 | 5.76 |

Patient name: Mr.Nandkumar Yadav

| Parameters | Apple lab Results | Aspira Lab Results | Nucleus Lab Results | CV% |
|------------|-------------------|--------------------|---------------------|------|
| Hb | 14.4 | 14.6 | 14.4 | 0.79 |
| PCV | 44 | 44.6 | 44.2 | 0.69 |
| RBC | 5 | 5 | 5.03 | 0.34 |

Patient name :Mrs.Indira Shroff

| Parameters | Apple lab Results | Aspira Lab Results | Nucleus Lab Results | CV% |
|------------|-------------------|--------------------|---------------------|------|
| Hb | 12.1 | 12.3 | 12.2 | 0.81 |
| PCV | 37.7 | 39.9 | 37.8 | 3.22 |
| RBC | 4 | 4.11 | 4.05 | 1.35 |
| ESR | 25 | 25 | 26 | 2.27 |

Patient name:Mr.Rajesh Panchal

| Parameters | Apple lab Results | Aspira Lab Results | Nucleus Lab Results | CV% |
|------------|-------------------|--------------------|---------------------|------|
| Hb | 13.5 | 13.6 | 13.3 | 0.52 |
| PCV | 40.9 | 43.2 | 42.4 | 2.76 |
| RBC | 4.5 | 4.61 | 4.6 | 1.26 |

Note:

Interlab results are acceptable.

Done By

Verify By QM

Approved By (Lab Director)