

CERTIFICATE OF ACCEPTABLE PERFORMANCE

Arpan Diagnostic Centre
 Biochemistry
 150ft Road, Indira Circle
 Nr Raj Bank
 RAJKOT
 360005
 INDIA

LABORATORY REF. NO. I91338/D

MONTHLY CLINICAL CHEMISTRY - CYCLE 16

30/12/2019

This is to certify that the above participant took part in a cycle of external quality assessment and achieved an acceptable level of performance (Cycle Average Absolute SDI ≤ 2) for the following parameters:

| | Cycle Average Absolute SDI |
|---|----------------------------|
| Albumin - Bromocresol Green - Erba EM Series | 0.89 |
| Alkaline Phosphatase - AMP optimised to IFCC - Erba EM Series | 0.36 |
| ALT (GPT) - Tris buffer without P5P - Erba EM Series | 1.42 |
| Amylase, Total - Other non blocked pNPG7 - Erba EM Series | 0.94 |
| AST (GOT) - Tris buffer without P5P - Erba EM Series | 0.55 |
| Bilirubin, Direct - Diazo with Sulphanilic Acid - Erba EM Series | 1.49 |
| Bilirubin, Total - Diazo with Sulphanilic Acid - Erba EM Series | 0.47 |
| Calcium, Ionised - Ion Selective Electrode - Medica Easylyte instruments | 0.26 |
| Cholesterol - Cholesterol Oxidase - Abell Kendall - Erba EM Series | 1.17 |
| CK, Total - CK-NAC (IFCC) - Erba EM Series | 1.49 |
| Creatinine - Other enzymatic methods - Erba EM Series | 0.68 |
| GGT - Gamma glut'3-carb'4-nitro(IFCC) - Erba EM Series | 0.96 |
| Glucose - Glucose oxidase - Erba EM Series | 1.21 |
| HDL-Cholesterol - Direct HDL, PEGME - Erba EM Series | 1.03 |
| Phosphate, Inorganic - Phosphomolybdate UV - Erba XL Series | 1.10 |
| Potassium - ISE method - direct - Medica Easylyte instruments | 1.24 |
| Protein, Total - Biuret reaction, end point - Erba EM Series | 1.01 |
| Sodium - ISE method - direct - Medica Easylyte instruments | 1.86 |
| Total T4 - Beckman Access/LXi725 - Beckman Access Series | 1.85 |
| Trig Total - Lipase/GPO-PAP no correction - Erba EM Series | 1.03 |
| Urea - Urease, kinetic - Erba EM Series | 0.75 |
| Uric Acid (Urate) - Uricase Perox. with ascorb. ox @ 546nm - Erba EM Series | 1.39 |

This certificate is valid until 24/01/2021

Authorised by:
 Stephen Doherty, RIQAS Manager

Page 1 of 1