

2/11/2020

External Quality Assurance Scheme - Print Monthly Summary

E

LAB MONTHLY SUMMARY

Lab Name LABORATORY DEPARTMENT, GENERAL HOSPITAL, BHARUCH Lab No 11916
 Month January Year 2020
 Constituent Group Chemistry II

Date of Result Entered : 18/01/2020

Date of Report Published : 11/02/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	UV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	OTHERS (Any other Analyzer)	850	134.43	9.89	13.03	128.5 mg/dl	-0.48	0.89
2	UREA I	UREASE BERTHELOT	OTHERS (Any other Analyzer)	281	39.56	14.36	5.68	40.9 mg/dl	0.24	0.88
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	OTHERS (Any other Analyzer)	465	1.78	9.53	0.17	1.9 mg/dl	0.71	0.02
4	T-PROTEIN I	BIURET - colorimetric	OTHERS (Any other Analyzer)	690	5.22	11.48	0.60	5.2 g/dl	-0.03	0.05
5	CHOLESTEROL I	CHOD-PAP	OTHERS (Any other Analyzer)	776	100.70	13.79	13.88	75.9 mg/dl	-1.79	1.00
6	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	OTHERS (Any other Analyzer)	766	115.15	15.91	18.32	100.5 mg/dl	-0.80	1.32

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****