



Ref No. TS/SBP/9251

Date: 1st January' 2020

Dr.Dharap's Diagnostic Centre
1st Floor, Limaye Niwas, Baburao Parulekar Rd,
Dadar West, Mumbai, Maharashtra 400028

Dear Sir,

As desired by you please find herewith the Calibration report

1. Year of manufacturing : EC-5 V2 Sr No. N150257
2. Range of the instrument : Optical density Linear form 0-2.5
3. Date of Calibration - 1st Januray' 2020

Test Voltages / Test Temperature -

Test Voltage	Actual Voltage	Ok / Not Ok
Logic 5 V	5V	OK
Lamp Voltage (11.5V – 12V)	11.72V	OK
SMPS V (17-19 V)	17.9V	OK
Flowcell Temperature	25 ⁰ C, 30 ⁰ C, 37 ⁰ C , OFF	ALL OK
Pump Calibration		OK

.....2



- 2 -

Sr.No.	Filters	Pre Channel	Log Channel
1	340	32764	0.7926
2	405	69394	0.4577
3	450	98494	0.3074
4	505	98085	0.3088
5	546	73199	0.4353
6	578	104721	0.2816
7	630	91083	0.3412
8	670	88450	0.3545

Thanking you and assuring of our prompt attention at all times.

Yours faithfully,

For TRANSASIA BIO-MEDICALS LTD.



REGIONAL SERVICE MANAGER