

## VIEW LAB MONTHLY SUMMARY

Lab Name	BALAJI MEDICAL CENTRE	Details About Robust Analysis								
Lab No	11452	Detail About Monthly Summary								
Month	January	Detail about SDI								
Year	2020									
Constituent Group	Chemistry II									
Click on the analyte to view Graphical Data										
<input type="button" value="All Method Result"/> <input type="button" value="Print"/>										
Date of Result Entered : 13/01/2020										
Date of Report Published : 11/02/2020										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE I	GOD-POD	OTHERS ( Any other Analyzer )	850	134.43	9.69	13.03	138 mg/dl	0.27	0.89
2	UREA I	UREASE BERTHELOT	OTHERS ( Any other Analyzer )	281	39.56	14.36	5.68	44.6 mg/dl	0.89	0.68
3	T.BILIRUBIN I	DIAZONIUM SALT ( Colorimetric) /JENDRASSIK	OTHERS ( Any other Analyzer )	597	2.04	21.38	0.44	2.4 mg/dl	0.83	0.04
4	T-PROTEIN I	BIURET - colorimetric	OTHERS ( Any other Analyzer )	690	5.22	11.48	0.60	5.2 g/dl	-0.03	0.05
5	ALBUMIN I	BCG - colorimetric	OTHERS ( Any other Analyzer )	674	3.19	10.36	0.33	3.1 g/dl	-0.27	0.03
6	URIC ACID I	ENZYMATIC / URICASE Colorimetric	OTHERS ( Any other Analyzer )	631	5.96	15.93	0.95	6.2 mg/dl	0.25	0.08
7	CHOLESTEROL I	CHOD-PAP	OTHERS ( Any other Analyzer )	776	100.70	13.79	13.88	145.1 mg/dl	3.20	1.00
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	OTHERS ( Any other Analyzer )	766	115.15	15.91	18.32	122 mg/dl	0.37	1.32
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	BALAJI MEDICAL CENTRE		<b>Details About Robust Analysis</b>							
<b>Lab No</b>	11452		<b>Detail About Monthly Summary</b>							
<b>Month</b>	January		<b>Detail about SDI</b>							
<b>Year</b>	2020									
<b>Constituent Group</b>	HbA1c									
Click on the analyte to view Graphical Data			<input type="button" value="All Method Result"/>		<input type="button" value="Print"/>					
			Date of Result Entered : 13/01/2020							
			Date of Report Published : 11/02/2020							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	TURBIDIMETRIC INHIBITION IMMUNO ASSAY	OTHERS ( Any other Analyzer )	171	5.33	11.42	0.61	4.2 %	-1.86	0.09
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>