



**Jeevan Healthcare**

Services Private Limited (OPC)

CALIBRATION CERTIFICATE

CUSTOMER : BALAJI MEDICAL CENTER  
ADDRESS : No-39-11-63/4-1, MURALI NAGAR,  
VISAKHAPATNAM ,ANDHRA PRADESH - 530007.

REPORT NO : 1511  
DATE OF CALIBRATION : 21-10-2019

CERTIFICATE NO : 1511  
NEXT CALIBRATION : 20-10-2020

MODEL : DIATRON  
NAME OF EQUIPMENT : ICOUNT5 Hematology analyzer SR-NO : 550034

CALIBRATION DONE BY : MR.G.V.KARTHIK

PARAMETER	RANGE	ACCURACY
WBC.	0.1-99.9 X 10 <sup>3</sup> / $\mu$ l.	3%
RBC.	0.2-7.0 X 10 <sup>6</sup> / $\mu$ l.	2%
HGB.	0.1 - 25.0 g/dL.	2%
HCT.	$\leq$ 0.5 %.	3 %
PLT	1-999 X 10 <sup>3</sup> / $\mu$ l	5%

We hereby the above equipment was calibrated and instrument was in good condition.



FOR JEEVAN HEALTHCARE SERVICES PVT LTD (OPC)

Chennai  
600 083

G.V.KARTHIK  
SERVICE MANAGER