

Calibration Certificate

UNIT DESCRIPTION:	RESEARCH CENTRIFUGE	MODEL NO.:	PR - 24
SR. NO.:	ZBLN-35650	CERTIFICATE NO.:	RSM - 0181/20
NAME OF CUSTOMER:	GENERAL HOSPITAL,		
ADDRESS OF CUSTOMER:	BHARUCH, GUJRAT		
CALIBRATION DATE:	02.01.2020	CAL. VALIDITY:	01.01.2021

DETAILS OF EQUIPMENT USED FOR CALIBRATION
 NOTE: Tick mark where as applicable.

Digital Lab Thermometer	<input type="checkbox"/>	Digital Tachometer	<input checked="" type="checkbox"/>	Digital Stop Watch	<input type="checkbox"/>
Auto Ranging R.P.M. Meter	<input type="checkbox"/>	Digital Strobe	<input type="checkbox"/>	Weight Set	<input type="checkbox"/>
Digital Temperature Indicator	<input type="checkbox"/>				

Sr. No.	Description	Certified By	Instrument Id. No.	Certificate No.	Calibration Date	Validity
01	Digital Tachometer	Autocal	TAC-02	C1960289/P/02	19.06.2019	18.06.2020

OBSERVATIONS:

Sr. No.	Instrument Speed Displayed (RPM)	Speed Reading on Digital Tachometer	Diff. In Actual and Digital Tachometer	Remarks
01	1000	1002	02	OK
02	3000	3007	07	OK
03	7000	7048	48	OK

- NOTE: 1. Calibration Traceable to national standards.
 2. Overall deviation is within specified accuracy.
 3. Result is related only to the item calibrated.
 4. Maximum tolerance permissible is $\pm 1^\circ\text{C}$ & $\pm 5^\circ\text{C}$ whichever is applicable and RPM $\pm 2\%$ FSD and $\pm 5\text{gm}$ to set weight.

REMARKS: ACCEPTED



CALIBRATED BY: RAJESH TIWARI

APPROVED BY: MANAGER Q. A.