

## Calibration Certificate

UNIT DESCRIPTION: ROTARY SHAKER MODEL NO.: RS 12 R  
 SR. NO.: ZEJS-36013 CERTIFICATE NO: RSM - 1551/19  
 NAME OF CUSTOMER: GENERAL HOSPITAL,  
 ADDRESS OF CUSTOMER: BHARUCH, GUJARAT  
 CALIBRATION DATE: 22.10.2019 CAL. VALIDITY: 21.10.2020

DETAILS OF EQUIPMENT USED FOR CALIBRATION: Digital Stop Watch

### DIGITAL STOPWATCH CALIBRATION TRACEABILITY DETAILS:

Sr. No.	Description	Certified By	Instrument Id. No.	Certificate No.	Calibration Date	Validity
01	Digi. Stopwatch	AUTOCAL	SW-104	C1980137/E/01	19.08.2019	18.08.2020

### OBSERVATIONS:

Sr. No.	Display Readings per minute	Speed Reading by Stop Watch per minute	Difference between Display & Stop watch	Remarks
01	180	180	00	OK

REMARKS: ACCEPTED

NOTE : 1. Calibration Traceable to national standards.  
 2. Overall deviation is within specified accuracy.  
 3. Result is related only to the item calibrated.  
 4. Maximum tolerance permissible is  $\pm 1^{\circ}\text{C}$  &  $\pm 5^{\circ}\text{C}$  whichever is applicable and RPM  $\pm 2\%$  FSD and  $\pm 5\text{gm}$  to set weight.



CALIBRATED BY: RAJESH TIWARI

APPROVED BY: MANAGER Q. A.