



# A to Z Biomedical Services

426, 4th Floor, pramukh TANGENT, KH-0, Sargasan Circle, KH-0 to GIFT City Road, Gandhinagar-382421, Gujarat  
Mo: +91 - 7874807758 | e-mail: atoz\_bm@yahoo.com

## CALIBRATION REPORT

Details of Device Under Calibration :

Date: 05-08-2019

Customer Name & Address: General Hospital, Bharuch	Date of Issue : 05-08-2019
Name of Equipment: Hot Air Oven	Calibration Date : 01-08-2019
Manufacturer : Labotech	Calibration Due Date : 31-07-2020
Model/SN. : BDI-50/200151	Location of equipment :
Certificate ID : AZ/19-20/ 1009	Environment Condition Temperature: 29° C
	Humidity: 35%

### Test Equipment Used:

Sr no	Test equipment	Company	Model	Sr no.	Cali. Date	Due date
1	Temp. Indicator with RTD sensor	Masibus	408	1002ME11278	17-01-19	16-01-20

### Test Report:

Condition of Equipment at the time of Calibration: Good

Sr No	Test	Unit	Set Value	Observed Value	Error	Acceptable Range (+/-)	Result
1	Temperature	° C	93	94	-1	5%	Pass

### Remarks:

- (1) Results reported are valid at the time of and under the stated conditions of measurement.
- (2) The Calibration due date is decided by user, However, the due date shall be recommended based on the customer's written advice.
- (3) Calibration has been done as per customer's requirement.
- (4) Measurements are traceable to the National Standard.
- (5) This document shall not be reproduced, except in full, without the written approval of "A to Z Biomedical Services"
- (6) DUC means Device Under Calibration, NS means Non Specified.
- (7) The Calibration was performed by comparison of DUC's Values ( Observed Values) and Test Equipment's Values ( Set Value).

Calibrated By:  Anal Patel	Approved By:  Gaurav Desai	
----------------------------------	----------------------------------	--