

CALIBRATION REPORT

Date: 05-08-19

Details of Device Under Calibration :

Customer Name & Address: General Hoospital, Bharuch	Certificate ID : AZ/19-20/ 1012A
	Date of Issue : 05-08-19
Name of Equipment: Domestic refrigerator	Calibration Date : 01-08-19
Manufacturer : Godrej	Calibration Due Date : 31-07-20
Model : F-3	Location of equipment : _____
Serial No : NS	Ambient Temperature : 22 °C

Test Equipment Used:

sr no	Test equipment	Company	Model	Sr no.	Cali. date	Cali. due date
1	Temperature Indicator with RTD sensor	Masibus	408	1002ME11278	17-01-19	16-01-20




Condition Of Equipment at the time of Calibration : GOOD

Test Report:

Sr No	Test	Unit	Range	Reading at	Observed Value by Test equipment	Result
1	Temperature	Deg.° C	2°C to 8 °C	Rack 1	4.4	Pass
2				Rack 2	4.5	Pass
3				Container	4.5	Pass
4			N.S.	Freezer	-16.7	Pass

Note:

- Results reported are valid at the time of and under the stated conditions of measurement.
- Calibration has been done as per customer's requirement.
- Measurements are traceable to the National Standard
- NS means Not Specified and DUC means Device Under Calibration
- The reported measurement uncertainty is estimated at a level of confidence of approx 95 % with a coverage factor $k = 2$.
- Cali means calibration .

Calibrated By:  Anal Patel	Approved By:  Gaurav Desai	
---	---	---

THIS DOCUMENT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVAL OF " A TO Z BIOMEDICAL SERVICES"