

CALIBRATION REPORT

Details of Device Under Calibration :

Date: 05-08-2019

Customer Name & Address: General Hospital, Bharuch		Date of Issue : 05-08-2019
Name of Equipment: <u>Deep Freezer</u>		Calibration Date : 01-08-2019
Manufacturer : Labotech	Model/SN. : LT825/20190453	Calibration Due Date : 31-07-2020
Certificate ID : AZ/19-20/ 1017		Location of equipment : _____
		Environment Condition Temperature: 29° C
		Humidity: 35%

Test Equipment Used:

sr no	Test equipment	Company	Model	Sr no.	Cali. Date	Due date
1	Temp. Indicator with RTD sensor	Masibus	408	1002ME11278	17-01-19	16-01-20




Test Report:

Condition of Equipment at the time of Calibration: Good

Sr No	Test	Unit	Set Value	Observed Value	Error	Acceptable Range (+/-)	Result
1	Temperature	° C	-20	-19.9	-0.1	5%	Pass

Remarks:

- (1) Results reported are valid at the time of and under the stated conditions of measurement.
- (2) The Calibration due date is decided by user, However, the due date shall be recommended based on the customer's written advice.
- (3) Calibration has been done as per customer's requirement.
- (4) Measurements are traceable to the National Standard.
- (5) This document shall not be reproduced, except in full, without the written approval of "A to Z Biomedical Services"
- (6) DUC means Device Under Calibration, NS means Non Specified.
- (7) The Calibration was performed by comparison of DUC's Values (Observed Values) and Test Equipment's Values (Set Value).

Calibrated By:  Anal Patel	Approved By:  Gaurav Desai	
---	---	---