

Preventive Maintenance Report

Date: 25/09/19

Customer's Name & Address: General Hospital
Bharuch

Concerned Person : _____ Department _____

Email Id : _____

Telephone : _____

Model / Part No.	Accessories / Attachments	Serial No.
MLX	Fine & Focusing Adjustment Working Ok Sample Holder Working Ok able To hold Sample Nosepiece Work Smoothly and Stop at Click Condenser Up/Down, Work Smoothly and stop At Click Halogen Bulb Not Flickering, Light Intensity Control Able To Change Intercity Of Light Eyepiece & Objectives Clean	16E0462

Serviced BY: Mohit S Pawar
 Mohit.s.Pawar

REMARKS IF ANY: _____

Customer's Signature with Stamp

Dr. Jyoti N. Vasava
 Pathologist (CI-1)
 General Hospital, Bharuch.