
	Shraddhdeep Greencross pathology laboratory, Sector-7	
	EQAS- corrective action form	


Outlier Test parameter	Bilirubin - Total, Indirect
Program, Cycle,	Cycle - 18, 13/01/2020
Z score, peer value	Total Bil → -3.17, Bili In → -4.36
Date of analysis	21/01/2020
Analysis done by (technician)	Mr. Maulik Gajjar
Date of received, condition	24/01/2020, Satisfactory
Reconstitution process	Satisfactory
Delay in analysis if any	No delay
Clerical error- wrong result, wrong method	Entry result OK, Method OK
Instrument function checks -Temp, blank	Satisfactory
Maintenance	Satisfactory
Instrument Calibration	OK
Reagent issues	No issues of reagent
Instrument probe	OK
Manufacturer problem-reagent/calibrator	NO
Carry over from previous sample	NO
Automated pipette precision	Satisfactory
IQC trend, analysis ICV1	Satisfactory
Parameter calibration	Satisfactory
PT provider -sample issue	NO ISSUES
PT provider-statistics issue	Number of Lab, statistics satisfactory

CML Vellore EQAS Z score satisfactory in Jan 2020

Final conclusion

Random Error. Follow up the parameter.

Done by- Ms Puja Patel



Verify by - Dr Dinesh Rathod

