

To Whom It May Concern

This is to certify that the Fully Automated CLIA Processor – CL 900i (Sr. No. BB1-930001560) which is installed at **Leads Healthcare - Kalpetta** has been calibrated on 10/01/2020, as per the prevailing manufacturer's protocol.

The mechanical modules are checked for alignment and corrected for any deviation using specialized tools and as per manufacturer standards.

The back ground checks are completed and found to be well with in range.

Record the Effect Check Performed

<u>No.</u>	<u>Reading</u>	<u>Range</u>	<u>Remarks</u>
1	3988	2500-6000	OK
2	3946	2500-6000	OK
3	3912	2500-6000	OK

The calibration is done by service team of Avantor Performance Materials India Ltd., who is the sole distributor of this product in India.

The next calibration is due on 10/01/2021.

For **Avantor Performance Materials India Ltd.**



Authorised Signatory