

## LAB MONTHLY SUMMARY

Lab Name S S HOSPITAL  
 Month January  
 Constituent Group Chemistry II

Lab No 7698  
 Year 2020

Date of Result Entered : 18/01/2020

Date of Report Published : 11/02/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Robonik ( Semi automatic / Fully Automatic )	310	134.03	10.92	14.64	124 mg/dl	-0.69	1.66
2	UREA I	UREASE BERTHELOT	Robonik ( Semi automatic / Fully Automatic )	165	40.17	16.30	6.55	46 mg/dl	0.89	1.02
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Robonik ( Semi automatic / Fully Automatic )	136	1.79	9.73	0.17	1.9 mg/dl	0.63	0.03
4	T.BILIRUBIN I	DIAZONIUM SALT ( Colorimetric) /JENDRASSIK	Robonik ( Semi automatic / Fully Automatic )	139	2.02	21.17	0.43	2.3 mg/dl	0.65	0.07
5	T-PROTEIN I	BIURET - colorimetric	Robonik ( Semi automatic / Fully Automatic )	154	5.29	9.84	0.52	5.6 g/dl	0.60	0.08
6	ALBUMIN I	BCG - colorimetric	Robonik ( Semi automatic / Fully Automatic )	142	3.25	12.95	0.42	3.2 g/dl	-0.12	0.07
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Robonik ( Semi automatic / Fully Automatic )	171	5.93	17.66	1.05	8 mg/dl	1.98	0.16
8	CHOLESTEROL I	CHOD-PAP	Robonik ( Semi automatic / Fully Automatic )	300	105.37	12.77	13.46	111 mg/dl	0.42	1.55
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Robonik ( Semi automatic / Fully Automatic )	166	116.97	21.29	24.91	118 mg/dl	0.04	3.87

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*