

Analyte	Mean for Comparison	Your Result	SDI	RMSDI	%DEV	RM%DEV	TS	RMTS	Performance
Haemoglobin	14.431	14.600	0.47	Too Few	1.2	Too Few	105	Too Few	
Haematocrit (HCT)	41.936	44.000	1.33	Too Few	4.9	Too Few	59	Too Few	
HbA1c	29.367	28.800	-0.62	Too Few	-1.9	Too Few	100	Too Few	
HbA1c	34.488	33.100	-1.08	Too Few	-4.0	Too Few	68	Too Few	
HbA1c	85.068	87.000	0.73	Too Few	2.3	Too Few	93	Too Few	
Mean Platelet Volume	9.813	9.100	-1.11	Too Few	-7.3	Too Few	67	Too Few	
Platelets (Impedance Count)	459.265	447.000	-0.35	Too Few	-2.7	Too Few	118	Too Few	
WBC (Impedance Count)	4.934	5.060	1.17	Too Few	2.6	Too Few	65	Too Few	
Red Cell Dist. Width CV	13.792	13.600	-0.24	Too Few	-1.4	Too Few	120	Too Few	
Red Cell Dist. Width SD	46.850	52.100	1.57	Too Few	<u>11.2</u>	Too Few	<u>49</u>	Too Few	
WBC (Impedance Count)	19.218	19.250	0.04	Too Few	0.2	Too Few	120	Too Few	

ORMSDI N/A

ORM%DEV N/A

ORMTS N/A

