



QAF-EQASML  
CYCLE SUMMARY REPORT  
CYCLE 25, Oct - 2019  
QAF-EQAS-ML-Hematology



<b>Lab ID</b>	3813	<b>Lab Name</b>	SHRADDHADEEP GREENCROSS Pathology Laboratory
<b>Branch</b>	Gandhinagar	<b>Release Date</b>	16/10/2019

Sr. No.	Test Name	Test Detail	Your Result	Target Value	SD	SDI	Comment	No.Of Lab	Target Rule
1	<a href="#">ABO Blood Group</a>	<b>Method</b> Column Agglutination by Gel Card <b>Instrument</b> Tulip MATRIX Gel System <b>Reagent</b> Tulip Diagnostics <b>Unit</b> Not Applicable	<b>Result</b> A <b>Grading/Titre</b> <b>Pattern</b> <b>Interpretation</b>	A	NA	NA	Satisfactory	9	method group
2	<a href="#">D-Dimer</a>	<b>Method</b> ELFA <b>Instrument</b> Biomerieux Mini Vidas/ Vidas <b>Reagent</b> Biomerieux <b>Unit</b> ng/mL	<b>Result</b> 167.5 <b>Grading/Titre</b> <b>Pattern</b> <b>Interpretation</b>	150.08	24.34	0.7156	Satisfactory	9	Peer Group Mean
3	<a href="#">Malarial Antigen</a>	<b>Method</b> Immunochromatography <b>Instrument</b> Not applicable/ Manual <b>Reagent</b> MERIL DIAGNOSTICS <b>Unit</b> Not Applicable	<b>Result</b> Negative <b>Grading/Titre</b> <b>Pattern</b> <b>Interpretation</b>	Negative	NA	NA	Satisfactory	59	method group
4	<a href="#">Rh Blood Group</a>	<b>Method</b> Tube Agglutination <b>Instrument</b> Not applicable/ Manual <b>Reagent</b> Tulip Diagnostics <b>Unit</b> Not Applicable	<b>Result</b> RhD Positive <b>Grading/Titre</b> <b>Pattern</b> <b>Interpretation</b>	RhD Positive	NA	NA	Satisfactory	34	method group
5	<a href="#">DCT</a>	<b>Method</b> Column Agglutination by Gel Card <b>Instrument</b> Tulip MATRIX Gel System <b>Reagent</b> Tulip Diagnostics <b>Unit</b> Text & Titre	<b>Result</b> Negative <b>Grading/Titre</b> <b>Pattern</b> <b>Interpretation</b>	Negative	NA	NA	Satisfactory	17	method group

Notes

EQAS PROVIDER  
Quality Assurance Forum,  
(Under the aegis of GAPM)  
Gujarat Association Of Pathologist and Microbiologist

Reviewed & Approved By  
Chairman QAF-EQASML