

CLINI CARE PATHOLOGY LABORATORY

Form No-

*For Creatinine
Dec-2019 cycle*

EQAS outlier corrective action check list

CMC Vellore

Sr No	ERRORS WITH PROFICIENCY TESTING	Check list
i.	Clerical error	
1.	Transcription error (may be pre- or post-analytical).	No
2.	Situations where wrong method has been registered for analysis or method change not updated	No
ii.	Methodological problem	
1.	Instrument function checks (e.g., temperatures, blank readings, pressures) not performed as necessary, or results not within acceptable range.	No
2.	Scheduled instrument maintenance not performed appropriately	No
3.	Incorrect instrument calibration.	No
4.	Standards or reagents improperly reconstituted and stored, or inadvertently used beyond expiration date.	No
5.	Instrument probes misaligned	No
6.	Problem with instrument data processing functions .The laboratory may need to contact the manufacturer to evaluate such problems.	No
7.	Problem in manufacture of reagents/standards, or with instrument settings specified by manufacturer	No
8.	Carry-over from previous specimen	No
9.	Automatic pipette not calibrated to acceptable precision and accuracy.	No
10.	Imprecision from result being close to detection limit of method.	No
11.	Instrument problem not detected by quality control: - QC material not run within expiration date, or improperly stored. - QC materials not run at relevant analyte concentration.	No
12.	Result not within range of reportable range (linearity) for instrument/reagent system.	No
13.	Obstruction of instrument tubing/orifice by clot or protein.	No
14.	Incorrect incubation times.	No
iii.	Technical problem	
1.	EQA material improperly reconstituted.	No
2.	Testing delayed after reconstitution of EQA material (with problem from evaporation or deterioration).	No
3.	Sample not placed in proper order on instrument.	No
4.	Result released despite unacceptable QC data.	No
5.	QC data within acceptable limits, but showed trend suggestive of problem with the assay.	No
6.	Inappropriate quality control limits/rules. If the acceptable QC range is too wide, the probability increases that a result will fall within the acceptable QC range yet exceed acceptable limits for EQA.	No
7.	Manual pipetting/diluting performed inaccurately, at an incorrect temperature, or	No

Reviewed by _____

-17/12/19

