

Laboratory department,

S.S.Hospital, Petlad.

Date: 20/02/2020

To,

Lab director,

Sterling Accuris Diagnostic,

Ahmadabad.

Sub: To sign MOU for inter-laboratory comparison program

Respected sir,

We laboratory department ,S.S.Hospital,Petlad have applied for Quality Assurance Scheme for Basic Composite Medical Laboratories and as per NABL requirement we need inter-laboratory comparison program and we wish to sign MOU WITH YOU FOR below mentioned test for same.

Kindly do needful and oblige

Thanking you

List of test:

CBC

Blood group

ESR

Prothrombin time

P/S Examination

Urine routine and micro

UPT

Stool R/M

Your faithfully



Pathologist Class-I  
S. S. Hospital, Petlad,  
Di. Anand

Accepted.



Sterling Accuris Diagnostic, Ahmad





**LABORATORY REPORT**

Name : M B MENGAR 1483  
 Lab ID : 2002113753  
 Sex/Age : F / 40 Years  
 Ref. By :  
 Location :

Registration on : 21-Feb-2020  
 Collected on : 21-Feb-2020 16:33  
 Approved on : 21-Feb-2020 17:11  
 Sample Type : URINE

TEST	RESULTS	UNIT	BIOLOGICAL REFERENCE RANGE
------	---------	------	----------------------------

**PREGNANCY TEST (UPT)**  
*Immunochromatography*

**URINE PREGNANCY TEST (UPT)**  
 Negative

----- End Of Report -----

*Yash*

This is an Electronically Authenticated Report  
 • Denotes Test not in NABL Scope  
 # Referred test

Note (LL-Very low, L-Low, H-High, HH-Very High)

Approved by : **DR YASH SHAH**  
 M.D (Path)  
 (For Sterling Accuris Diagnostics)

CAP accredited: Sterling Accuris Diagnostics, Ahmedabad  
 22-Feb-2020 11:59 AM

ILC REPORT: LAB ID: 1482 URINE ROUTINE AND MICROSCOPY

SR. NO	PARAMETER (URINE ROUTINE MICROSCOPY )	REPROT LABORATORY DEPARTMENT S.S.HOSPITAL ,PETLAD	REPORT STERLING ACCURIS LABORATORY AHMEDABAD	REMARKS	
1	COLOR	YELLOW	Pale yellow	Matching	
2	PH	6.0	7.0	Matching	
3	Specific Gravity	1.020	1.010	Matching	
4	Protein	Nil	Negative	Matching	
5	Sugar	Nil	Negative	Matching	
6	Bile pigment	Absent	Negative	Matching	
7	Microscopy :	Pus cells :0-1/hpf Rbc : - /hpf Epithelial cells: 0-1/hpf Crystal :1-2 /hpf	Pus cells :occasional Rbc: nil /hpf Epithelial cells: occasional /hpf Crystal : nil	Matching	

ILC REPORT: LAB ID: 1483 UPT

SR. NO	PARAMETER (UPT)	REPROT LABORATORY DEPARTMENT S.S.HOSPITAL ,PETLAD	REPORT STERLING ACCURIS LABORATORY AHMEDABAD	REMARKS	
1	UPT	NEGATIVE	NEGATIVE	MATCHING	

Note: report of both laboratory attached.



Pathologist Class-I  
S. S. Hospital, Petlad,  
Di. Anand

**LABORATORY SERVICES OF S.S.H.  
PETLAD .**

<b>Reporting Form</b>	<b>Hematology &amp; Clinical pathology Report</b>	<b>Unique ID.LSGGGH/F/</b>
-----------------------	---	----------------------------

**S.S HOSPITAL,PETLAD  
Department of Pathology**

(Patient Sticker)

**Haematology & Clinical Pathology Reporting Form**

Patient Name: M. B. Menya

Age: 40 Yrs. Gender: M /  F Unit: \_\_\_\_\_

MRD No.: 12401 OPD Reg. No.: \_\_\_\_\_

Indoor Reg. No.: \_\_\_\_\_ Ward No. /OPD: \_\_\_\_\_

Head of Unit's Name: \_\_\_\_\_

Provisional Diagnosis: \_\_\_\_\_

Time of Sample Receiving: 9H/00M AM/PM

**Lab I. D. No.:** 1483

<b>Date of Report</b>	<u>21/02/2020</u>
<b>Time of Report</b>	<u>10:00 AM/PM</u>

HAEMATOLOGY			
Test	Result	Unit	Reference Range (Adult)
Hemoglobin		gm/dl	M: 14.0-17.0 gm/dl F: 12.0-16.0 gm/dl
Total WBC Count		/cmm	4000-10000/cmm
<b>Differential WBC Count:</b>			
Neutrophils		%	55-70%
Lymphocytes		%	20-40%
Eosinophils		%	1-6%
Monocytes		%	2-10%
Basophils		%	0-1%
RBC		10 <sup>12</sup> /l	4.0-5.50 10 <sup>12</sup> /l
HCT		%	36.0-52.0 %
MCV		fl	76.0-96.0 fl
MCH		pg	27.0-32.0 pg
MCHC		g/dl	30.0-35.0 g/dl
RDW		%	11.5-14.5 %
Platelet Count		lacs/cmm	1.5-4.5 lacs/cmm
ESR		mm/hr	M: ≤10 mm/hr F: ≤15 mm/hr
AEC		/cmm	40-440/cmm
Blood Group			

