

SERVICE / INSTALLATION / MAINTENANCE REPORT

GSTIN: 07BBPPS4039C1ZI

Sr. No. : NB 2708

Date / dates of visit : - 16-08-019 Time : 11:35 AM Total hours spent on job : 2 Hours

Customer's name - RMRIMS Phone : () - 0612-263665 Instrument : - CRT
 Customer's address - AUAMKUAN, PATNA07 Fax : () SI. Number : - NO1402160
 Contact person's name - MR. SUBHASH KR. e-mail Model : - NOVA-1

Purpose of visit : Service visit / Routine visit / Installation / Maintenance Service Contract / Customer Training Free Service / Inspection / Paid Service.

Problems reported by customer in case of service / maintenance visit :

Engineer's diagnosis
 FOR SERVICE CALL

Action taken by engineer
 DO RUN QC, Calibration Done.

Material used during visit :				Status of Instrument after engineer's visit :		QC performance for :				
Part No.	Quantity	Item	Utilized from	Remarks in any	M.Parameters	Slope	M.Parameters	L1	L2	L3
					pH		pH			
					pCO2		pCO2			
					pO2		pO2			
					SO2%		SO2%			
					Hct	11.2	Hct			
					HB		HB			
					Na	9.6	Na	147	166	
					K	8.8	K	4.0	5.7	
					Cl		Cl			
					iCa		iCa			
					iMg		iMg			
					Glucose		Glucose			
					Lactate		Lactate			
					BUN		BUN			
					Creat		Creat			
					TCO2		TCO2			
					LI		LI			
					HMB		HMB			
					O ₂ Hb		O ₂ Hb			
					MetHb		MetHb			
					COHb		COHb			
					tBILI		tBILI			

Engineer's suggestions / comments :
 All Parameter working satisfactory
 & QC Level-1 & Level-2 Run.

Date : - 16-08-019

Customer's comments : working satisfactory

[Signature]
 16-08-19

Job completed : Yes No
 If YES, it is certified that the instrument is working all respect

It is certified that the job as above has been certified out by your engineer to my / our satisfication
 The instrument is working in all respect.

Amit Kumar
 ENGINEER'S SIGNATURE & NAME Date 16-08-019

Customer's Name : - RMRIMS
 Department : - Central Diagnostic Lab
 Date : - 16-08-19
 Dear Customer, Please inform us about our overall after sales support.
 Excellent / Good / Better / Fair / Bad / Needs Improvement (Please cross whatever not applicable)

ST SURGITECH INDIA
 D-23, Flat No.-16, Rajouri Garden, New Delhi- 110 027 (India)
 Ph: +91-11-45525655 Fax: +91-11-41444846
 Mob. : +91-9910477777, +91-8560722268, +91-8800777774
 E-mail: info@surgitechindia.com

Exclusive Distributor of
NOVA
 biomedical
 U.S.A.