



## PROFICIENCY TESTING REPORT

### ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME

NABL accredited program as per ISO/IEC 17043:2010 standard  
Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No.: 285

Distribution No.: 147-A

Month/Year: April/2019

Instrument ID: Swelab Alfa (Sl. no. 16156)

Name & Contact No. of PT Co-ordinator: Dr. Renu Saxena (Prof & Head), Hematology, AIIMS, Delhi,  
Tel: 9013085730, E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 22-05-2019[Final].

#### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)					Within Lab (Precision Testing)		
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values	Z Score
WBC x10 <sup>3</sup> /μl	1	4.0	4.0	8	7.81	0.16	0	0.1	-1.5
RBC x10 <sup>6</sup> /μl	1	3.81	3.79	7.6	7.72	-0.74	0.02	0.03	-0.34
Hb g/dl	1	13.9	13.7	27.6	27.5	0.17	0.2	0.1	1.35
HCT%	1	42.5	42.4	84.9	87.1	-0.35	0.1	0.3	-0.54
MCV-fl	1	111.9	111.5	223.4	224.75	-0.1	0.4	0.2	0.45
MCH-Pg	1	36.4	36.3	72.7	71.35	0.73	0.1	0.2	-0.45
MCHC-g/dl	1	32.6	32.4	65	63.2	0.43	0.2	0.2	0
Plt. x10 <sup>3</sup> /μl	1	111	104	215	262	-0.83	7	4	0.81
Retic Count %	2	2.7	2.6	5.3	4.6	0.25	0.1	0.2	-0.45

#### P.S. Assessment

YOUR REPORT			CONSENSUS REPORT
DLC%	3	Nrbcs=01 per 100 WBC's , Poly=59/-/11 L=05, E=02, Mono/Promono=03 , B1=02 P.M.=03, Myc=10, Meta=05, Other=WBC's are markedly increased in no. with mostly mature cells and some immature cells of myeloid series, Occasional smear cells, Platelets adequate in no.	Poly: 60-70, Lymph: 2-6, nRBC/Eo/Mono/Blast/Pro: 0-5, Myelo: 6-12, Meta: 6-12, Baso: 2-5
RBC Morphology	3	Mild to moderate hypochromic cells, some anisocytosis with many microcytes and some microcytes, Occasional poikilocytosis & polychromasia with occas. crenated RBC's, no nucleated RBC's seen	Predominantly: Normocytic & Normochromic, Anisocytosis. Moderate: Microcytic. Mild: Hypochromic
Diagnosis	3	Chronic Myeloid Leukaemia	Chronic Myeloid Leukemia (Chronic Phase)

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /μl	1	450	407	90.42	91.4	4.67	5.65	4.91	2.95
RBC x10 <sup>6</sup> /μl	1	450	407	89.93	85.01	4.67	6.88	5.41	8.11
Hb g/dl	1	450	407	86.98	86	7.37	7.37	5.65	6.63
HCT%	1	450	407	94.59	87.47	4.42	6.39	0.98	6.14
MCV-fl	1	450	406	94.33	89.41	3.69	3.94	1.97	6.65
MCH-Pg	1	450	404	91.83	82.92	5.2	7.67	2.97	9.41
MCHC-g/dl	1	450	406	96.06	86.95	2.71	5.67	1.23	7.39
Plt. x10 <sup>3</sup> /μl	1	450	406	93.6	86.7	2.96	8.13	3.45	5.17
ReticCount%	2	450	348	89.08	85.92	4.31	2.59	3.74	8.62
PS Assessment	3	450	408	Acceptable:96.8%,Warning Signal:3.2%,Unacceptable :0%					

**Comments:**

- 1). Among Lab (EQA) : Acceptable Results
- 2). Within Lab (IQA) : Acceptable Precision.

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values – Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values – Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between“0 to ±2” are texted in green colour. Z score value between“±2 to ±3” are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

Report authorized by,



Dr. R. Saxena  
 Prof & Head, Hematology, AIIMS, Delhi.  
 PT Co-ordinator: ISHTM-AIIMS-EQAP

-----End Of Report-----