

Lab Name BALAJI MEDICAL CENTRE Lab No 5652

Month January Year 2020

Group
Constituent HBA1c

Date of Result Entered : 20/01/2020

Date of Report Published : 11/02/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	HBA1c	ENZYMATIC	OTHERS (Any other Analyzer)	51	4.78	19.18	0.92	5 %	0.24	0.26

SDI Range		Interpretation
Within -1.0 to +1.0	Excellent.	
Between ±1.0 to ±2.0	Good.	
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.	
Beyond ±3.0	Unacceptable performance. Action Signal.	

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

Contact details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

***** End of Report *****