

**LABORATORY DEPARTMENT**  
**General Hospital, Bharuch**  
**Inter Lab Comparison Report of Laboratory Department &**  
**Government Medical College, Surat.**

TEST DONE ON 14/2/20

**HEMOGRAM**

ILC no.	LAB. no.	MRD NO	TEST	GHB	GMCS
1	11	2419	<b>HB</b>	11.8	11.6
2	84	16632		13.9	12.8
3	99	114621		9.4	9.4
4	111	16732		15.2	15.2
5	117	16371		8.8	8.6

1	11	2419	<b>RBC</b>	4.78	4.84
2	84	16632		4.84	4.43
3	99	114621		4.46	4.71
4	111	16732		4.80	4.98
5	117	16371		5.11	4.91

1	11	2419	<b>WBC</b>	14100	14100
2	84	16632		6800	5200
3	99	114621		10200	10,000
4	111	16732		10,400	10100
5	117	16371		6700	5800


1	11	2419	<b>PLATLET</b>	3.08	3.01
2	84	16632		2.64	2.31
3	99	114621		3.03	2.78
4	111	16732		2.34	2.24
5	117	16371		2.91	3.62

1	11	2419	<b>HCT</b>	36.9	34.2
2	84	16632		41.8	30.8
3	99	114621		30.3	29.2
4	111	16732		44.8	42.8
5	117	16371		30.5	24.2

1	11	2419	<b>MCV</b>	77	71
2	84	16632		87	70
3	99	114621		68	62
4	111	16732		93	85
5	117	16371		60	49

GHB=GENERAL HOSPITAL  
 BHARUCH

GMCS=GOVERNMENT MEDICAL  
 COLLEGE SURAT

  
 Lab director  
 Dr. Jyoti N. Vasava  
 Pathologist (CI-1)

ILC	LAB.	MRD	TEST		
no.	no.	NO		GHB	GMCS

1	11	2419	MCH	24.6	23.9
2	84	16632		28.7	28.9
3	99	114621		21	20
4	111	16732		31.7	30.6
5	117	16371		17.3	17.5

1	11	2419	RDW CV	13.5	15.6
2	84	16632		12.5	13.4
3	99	114621		14.4	17.1
4	111	16732		11.5	14.9
5	117	16371		16.9	16.3

				NEUT /LYMP	/EOSI /MONO
1	11	2419	DC	68/28/2/2	73/23/2/2
2	84	16632		50/44/2/4	63/32/3/2
3	99	114621		70/18/3/2	81/15/2/2
4	111	16732		58/35/4/3	56/40/2/2
5	117	16371		54/41/3/2	66/30/2/2

1	11	2419	PLATELET	ADEQUATE	ADEQUATE
2	84	16632		ADEQUATE	ADEQUATE
3	99	114621		ADEQUATE	ADEQUATE
4	111	16732		ADEQUATE	ADEQUATE
5	117	16371		ADEQUATE	ADEQUATE

  
Lab Director

*Dr. Jyoti N. Vasava*  
Pathologist (CI-1)  
General Hospital, Bharuch.