



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 3027

Distribution No.: 148-G

Month/Year: September/2019

Instrument ID: HORIBA YUMIZEN 500

Name & Contact No. of PT Co-ordinator: Dr. Renu Saxena (Prof & Head), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 19-11-2019[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /μl	1	1.47	1.2	2.67	9.65	0.0570	-5.00	0.27	0.1	0.0070	1.64
RBC x10 ⁶ /μl	1	4.9	4.9	9.8	9.72	0.0110	0.29	0	0.04	0.0030	-0.77
Hb g/dl	1	11.2	11.1	22.3	22.6	0.0290	-0.45	0.1	0.1	0.0070	0.00
HCT%	1	41.6	41	82.6	76.6	0.2280	1.08	0.6	0.3	0.0260	0.81
MCV-fl	1	85.2	85	170.2	157.9	0.4100	1.27	0.2	0.2	0.0190	0.00
MCH-Pg	1	22.9	22.4	45.3	46.5	0.0570	-0.90	0.5	0.2	0.0140	2.02
MCHC-g/dl	1	26.9	26.4	53.3	59.1	0.1830	-1.35	0.5	0.2	0.0170	1.01
Plt. x10 ³ /μl	1	207	195	402	435	2.65	-0.50	12	6	0.41	1.01
Retic %	2	4.5	4.2	8.7	10	0.18	-0.27	0.3	0.4	0.03	-0.19

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT
DLC%	3	Nrbcs=NIL , Poly=14 L=80, E=01, Mono/Promono=10 , B1=00 P.M.=03, Mye=01, Meta=00, Other=	Lymph: 85-95, nRBC/Poly/Eo/Mono/Blast/Pro/Mye/Meta: 0-5, Lymphocytes morphology resembling a lot with blasts morphology.
RBC Morphology	3	NORMAL	Predominantly: Normocytic Normochromic, Moderate: Microcytic, Mild: Hypochromic.
Diagnosis	3		Chronic Lymphocytic Leukemia [CLL] / Acute Leukemia. Adv: Immunophenotyping by flow cytometry.

EQAP Code No.:
3027

Distribution No.: 148-G Month/Year: September/2019

Instrument ID: HORIBA YUMIZEN
500**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 ³ /µl	1	350	262	86.64	84.35	5.73	5.73	6.87	8.02
RBC x10 ⁶ /µl	1	350	262	85.88	91.98	6.11	2.29	7.63	4.96
Hb g/dl	1	350	262	86.64	89.69	3.44	5.34	9.54	4.58
HCT%	1	350	262	91.6	88.93	3.44	4.2	4.58	6.49
MCV-fl	1	350	262	91.22	91.22	5.34	1.91	3.05	6.49
MCH-Pg	1	350	262	86.64	82.82	6.11	9.16	6.87	6.87
MCHC-g/dl	1	350	262	92.37	92.37	4.2	1.91	3.05	5.34
Plt. x10 ³ /µl	1	350	262	93.13	88.17	4.2	5.73	1.91	5.34
ReticCount%	2	350	239	92.89	94.98	5.02	3.77	2.51	1.67
PS Assessment	3	350	238	Acceptable:84.3%,Warning Signal:6.1%,Unacceptable :9.6%					

***Comments:**

1). **Among Lab (EQA) : PS Diagnosis not reported, Remaining results acceptable**

2). **Within Lab (IQA) : Precision acceptable.**

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Report authorized by,



Dr. R. Saxena

Prof & Head, Hematology, AIIMS, Delhi.

PT Co-ordinator: ISHTM-AIIMS-EQAP

-----End Of Report-----