

AUSSIN (ASH) BIOMED CALIBRATION LABORATORY

508- DEV PRIME, B/H DIVYA BHASKAR PRESS, OFF S.G ROAD, AHMEDABAD, GUJARAT.

Ph. +91 9925063777, +91 9978163777

E-mail: monil.bmed@gmail.com

www.ashbme.com



Calibrator machine is calibrated under NABL Laboratory.

CERTIFICATE

CERTIFICATE /ID NO: **ASH- 5081**

COMPANY / HOSPITAL: **GENERAL HOSPITAL MORBI.**

ADDRESS: **GANDHI CHOWK, OPP. TOWN HALL, RAVAPAR ROAD, MORBI, GUJARAT-363641**

DESCRIPTION OF INSTRUMENT: **CENTRIFUGE MACHINE**

Make: **REMI**

Model: **NA**

SR NO: **NA**

LOCATION: **LABORATORY**

TESTED BY: **RAHUL PATEL**

DATE OF CALIBRATION: **15/10/2019**

NEXT CALIBRATION DATE: **14/10/2020**

TEMPERATURE & RH: **22 deg C & 45%**

USE OF CALBRATOR: **RPM CALBRATOR**

USE OF CALBRATOR	MAKE	MODEL	SR NO
DIGITAL TECHO METER	LUTRON	DT 2234C	S015345

A: CALIBRATION DATA:

Standard RPM	UUC* RPM	Error RPM	Uncertainty $\pm 5\%$
1000	1015	15	$\pm 5\%$
2000	2025	25	$\pm 5\%$
3500	3485	-15	$\pm 5\%$

* Mean of Five readings

Note: The indicated Uncertainties are estimated expanded Uncertainties for Coverage factor 'k'=2 providing a level of confidence approx. = 95% for distribution

CALBRATED/TESTED BY: **RAHUL PATEL**

DATE: **17/10/2019**



The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement. This report is not to be used to claim product endorsement by Aussin any agency. This document shall not be reproduced, except in full, without the written approval of Aussin. This is for technical information of the client only and not for advertisement, promotion, publicity of litigation.