

AUSSIN (ASH) BIOMED CALIBRATION LABORATORY

508- DEV PRIME, B/H DIVYA BHASKAR PRESS, OFF S.G ROAD, AHMEDABAD, GUJARAT.

Ph. +91 9925063777, +91 9978163777

E-mail: monil.bmed@gmail.com

www.ashbme.com



Calibrator machine is calibrated under NABL Laboratory

CERTIFICATE

CERTIFICATE /ID NO: ASH- 5066

COMPANY / HOSPITAL: GENERAL HOSPITAL MORBI.

ADDRESS: GANDHI CHOWK, OPP. TOWN HALL, RAVAPAR ROAD, MORBI, GUJARAT-363641

DESCRIPTION OF INSTRUMENT: CENTRIFUGE MACHINE

Make: YORCO

Model: HUMLC-140

SR NO: NA

LOCATION: LABORATORY

TESTED BY: RAHUL PATEL

DATE OF CALIBRATION: 15/10/2019

NEXT CALIBRATION DATE: 14/10/2020

TEMPERATURE & RH: 22 deg C & 45%

USE OF CALIBRATOR: RPM CALIBRATOR

USE OF CALIBRATOR	MAKE	MODEL	SR NO
DIGITAL TECO METER	LUTRON	DT 2234C	S015345

A: CALIBRATION DATA:

Standard RPM	UUC* RPM	Error RPM	Uncertainty $\pm 5\%$
1000	1009	9	$\pm 5\%$
2000	1985	-15	$\pm 5\%$
3500	3530	30	$\pm 5\%$

* Mean of Five readings

Note: The indicated Uncertainties are estimated expanded Uncertainties for Coverage factor 'k'=2 providing a level of confidence approx. = 95% for distribution

CALIBRATED/TESTED BY: RAHUL PATEL

DATE: 17/10/2019



The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement. This report is not to be used to claim product endorsement by Aussin any agency. This document shall not be reproduced, except in full, without the written approval of Aussin. This is for technical information of the client only and not for advertisement, promotion, publicity of litigation.