

# AUSSIN (ASH) BIOMED CALIBRATION LABORATORY

508- DEV PRIME, B/H DIVYA BHASKAR PRESS, OFF S.G ROAD, AHMEDABAD, GUJARAT.

Ph. +91 9925063777, +91 9978163777

E-mail: [monil.bmed@gmail.com](mailto:monil.bmed@gmail.com)

[www.ashbme.com](http://www.ashbme.com)



Calibrator machine is calibrated under NABL Laboratory.

## CERTIFICATE

CERTIFICATE /ID NO: ASH-5069

COMPANY / HOSPITAL: GENERAL HOSPITAL MORBI.

ADDRESS: GANDHI CHOWK, OPP. TOWN HALL, RAVAPAR ROAD, MORBI, GUJARAT-363641

DESCRIPTION OF INSTRUMENT: PIPETTE

MAKE: I-PIPETTE

MODEL: NA

SR NO: NA

LOCATION: LABORATORY

TESTED BY: RAHUL PATEL

DATE OF CALIBRATION: 15/10/2019

NEXT CALIBRATION DATE: 14/10/2020

TEMPERATURE & RH: 22 deg C & 45%

USE OF CALIBRATOR: HIGH PRECISION E2 CLASS WEIGHTS

USE OF CALIBRATOR	MAKE	MODEL	SR NO
DIGITAL WEIGHING BALANCE	SF-400	NA	ASHW05

### A: CALIBRATION DATA:

Standard $\mu\text{l}$	UUC* $\mu\text{l}$	Error $\mu\text{l}$	Uncertainty $\mu\text{l} \pm 5\%$
10	10.009	0.009	$\pm 5\%$
30	29.987	-0.013	$\pm 5\%$
50	50.014	0.014	$\pm 5\%$
100	100.012	0.012	$\pm 5\%$

\* Mean of Five readings

Note: The indicated Uncertainties are estimated expanded Uncertainties for Coverage factor 'k'=2 providing a level of confidence approx.  $\approx 95\%$  for distribution

CALIBRATED/TESTED BY: RAHUL PATEL

DATE: 17/10/2019



The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement. This report is not to be used to claim product endorsement by Aussin any agency. This document shall not be reproduced, except in full, without the written approval of Aussin. This is for technical information of the client only and not for advertisement, promotion, publicity of litigation.