



Date: 28/11/2018

Certificate of Calibration

Customer Name: PATHKIND DIAGNOSTIC PVT LTD. GONDA

Model: Fully Automated Bio-Chemistry Analyzer EM-200

Serial No.: B200671

Calibration done date: 28/11/2018

Next Calibration due Date: 27/11/2019

*This is to certify that above mentioned product has been verified of calibration
By*

Checking Lamp Voltage 11.8V

Checking gain of Photo meter for all wavelength values obtained are as under

Filter	340	405	505	546	578	600	660	710
Gain	435	497	481	498	486	490	412	482

Checked calibration of probe on all locations.

*Checked temperature at Reagent tray at 8.0 degree (Range 4 to 11 degree) and
reaction tray 37.2 degree (Range 36.5 to 37.5 Degree)*

*Followed by calibration of all the three types of chemistries (picked up
CHOL, CREAT and SGPT) and QC run with satisfactory values of QC*

Calibration at site performed by Mr. SACHIN SINGH THAKUR

Sachin S.M.
Authorized Signatory
(Mr. SACHIN SINGH THAKUR)
Sr. SERVICE ENGINEER
For Transasia Bio-Medicals Ltd.