

## Calibration Certificate

This is to certify that the Cobas c111 is a fully automated analyzer with S.No:15351  
At Vijaya Diagnostic Pvt.Ltd, Nellore has been calibrated on 3<sup>rd</sup> November 2019.

The calibration includes:

- Replacement of Filters
  - Replacement of Lamp
  - Checked Voltages
  - Lubrication of Robotics
  - Adjustments and Alignments of Probe
  - Temperature Adjustments
  - Performed Air/Water Calibration
- 
- Reaction Disk: Range: 36.8°C to 37.2°C  
Adjusted: 37°C
  - Reagent Disk: Range: 8.0°C to 12.0°C  
Adjusted: 10°C

Adding to this all the temperatures and volumes drawn by pumps were checked and found they are OK.

Next calibration due on 2<sup>nd</sup> November 2020.

For Roche Diagnostics India Pvt Ltd.

Sathish Dasari,  
Technical Service specialist  
Vijayawada

