

**HORIBA India Private Ltd**



246, Okhla Industrial Estate, Phase - III  
110 020 New Delh, ilndia

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

**HIN/MED/2019-20/03218**

November 23 2019

## CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **ABX MICROS60** bearing serial no. **606OT96193** installed at **VIJAYA DIAGNOSTICS, NELLORE, ANDHRA PRADESH** was calibrated on **23 Nov 2019**.

Calibrator : **ABX MINOCAL**

Lot.No : **CX440**


Expiry Date : **05.12.2019**

The reports of blank, repeatability, calibration and quality controls were all found in accepted range.

Next Calibration Due: **November 22, 2020.**

**Ebin THOMAS**

(Product Manager)  
For Horiba India Pvt. Ltd.



Verified By  
*B. Manohar Babu*

