


LABORATORY SERVICES OF S.S.H.

Petlad

Formats	EQAS Outlier Form	Unique.ID:-LSSSH/
Doc No. LSSSH/		

EQAS / ILC	ILC	Month & Year	February - 2020
Parameter (Outlier)			
Corrective Actions Taken (Add yes/No in the Following Checklist & Explain in Remarks)			

Transcription Error	ILC NOT MATCHING		
QC Run on the day of EQAS Run			
Methodological Problems	Instrument Problem		
	Faulty Reagents		✓
	Incorrect Calibration		
	Other Methodological Problems		
Technical Problems	Time delay between EQAS Reconstitution & Actual Analysis		
	Incorrect Pipetting		
	Dilution Error		
	Calculation Error		
	Run Accepted in non-linear range		
	Run Accepted even though controls were out of range		✓
	Misinterpretation/Misidentification		
EQAS Material Problems	Other technical problems		
	Unstable PT Material		
	Improper transport		
	Contamination/Hemolyzed		
Corrective action taken	Other Problems		
	Reagent discarded New kit opened QC OK.		
Result after rerun	No issue		
Remarks & Sign	ILC result acceptable 		

Pathologist Class-I
 S. S. Hospital, Petlad,
 Di. Anand

Prepared By : Deputy Quality Manager	Version No. :	Page 1 of 1
Approved By : Lab. Director	Issue Date :	
Issued By :- Lab. Director	Amed. Date :	

LABORATORY REPORT



Name : KAMLESHBHAI C PATEL
 Lab ID : 2002801715
 Sex/Age : M / 41 Years
 Ref. By : DR. ANIL TRIVEDI
 Location : ASHWINI MEDICAL@ANAND

Registration on : 29-Feb-2020
 Collected on : 29-Feb-2020 18:05
 Approved on : 29-Feb-2020 19:36
 Sample Type : SERUM

TEST	RESULTS	UNIT	BIOLOGICAL REFERENCE RANGE
AMYLASE <small>Dyeid saccharidase based enzymatic colorimetric</small>	52.61	U/L	30 - 110
LIPASE <small>Colipase/Diacetone/DXG/PO/POD Two point specific method</small>	40.07	U/L	23 - 300

———— End Of Report ————

A.N. Trivedi

This is an Electronically Authenticated Report
 Genuine Test not to be sold, bought

Valid till 30th Dec 2020. For more details visit our website

Test done from collected sample

Page 1 of 1

Approved by : DR. ANIL TRIVEDI

Printed on : 02-Mar-2020 17:07

LABORATORY AT : Ashwini Medical Centre, B/h Kalpana Cinema, Off Bhalej Road, Anand - 388 001 Ph. 8849 883 805 | website: www.sterling-accuris.com

NATIONAL REFERENCE LABORATORY

FOR ILL

**LABORATORY SERVICES OF S.S.H.
PETLAD .**

Reporting Form	Clinical biochemistry Report	Unique ID.LSGGGH/F/
-----------------------	---	----------------------------

**S.S.Hospital, Petlad
Laboratory Department**

(Patient
Sticker)

OPD / IPD

Clinical Biochemistry Report Form

Patient Name: Kumlesh C Patel Lab I.D. No.: 05

Age: 41 Yrs. Gender: M / F Sample Receipt Time: 10:00 AM/PM Date: 01/08/2020

MRD No.: 15010 Result Reporting Time: 11:00 AM/PM Date: 01/08/2020

OPD Reg. No.: - Indoor Reg. No.: - Ward No. /OPD: 00 Unit: -

TEST	RESULT	UNIT	REFERENCE RANGE
DIABETIC PROFILE			
Plasma Glucose (Fasting)		mg / dl	70 – 110
Plasma Glucose (Post Prandial)		mg / dl	70 – 140
Plasma Glucose (Random)		mg / dl	70 – 140
HbA1c		%	< 6.5
RENAL FUNCTION TEST			
B1. Urea		mg / dl	5 – 40
S. Creatinine		mg / dl	0.6 – 1.4
LIVER FUNCTION TEST			
S. Bilirubin	Total	mg / dl	< 1
	Direct	mg / dl	< 0.3
	Indirect	mg / dl	0.0 – 0.7
ALT (SGPT)		U / L	5 – 42
AST (SGOT)		U / L	Up to 40
S. Alkaline Phosphates		U / L	38 – 126
S. Total Protein		gm / dl	6.6 – 8.5
S. Albumin		gm / dl	3.5 – 5.0
S. Globulins		gm / dl	2.3 – 3.5

